



## Calibration Certificate

Customer Name	Shri Mata Sheetla Devi Swarna Jayanti Diagnostic Centre Hindlabs HLL Lifecare Ltd. (A Govt. of India Enterprise)
Address	Sec-39, Jharsa, Gurugram -122003
Contact Person	Dr. Vijay Kumar Sharma
Contact No.	+91-8800336556
Instrument	Urine Dry Chemistry Analyzer
Model	Uri Plus-600
Serial No.	NI900500H0348
Instrument Status	Under Warranty
Calibration Done On	27.07.2021
Next Calibration Due On	26.07.2022

**For Aspen Diagnostics Pvt. Ltd.**

*for*  
*Hemant Singh Bisht*

Hemant Singh Bisht  
Technical Executive



*Rajaram Krishnamurthy*

Rajaram Krishnamurthy  
National Service Manager



# RAPID DIAGNOSTIC GROUP OF COMPANIES

\* Rapid Diagnostic Pvt. Ltd.      \* Aspen Diagnostics Pvt. Ltd.  
 \* Aspen Laboratories Pvt. Ltd.    \* Deep Meditech Pvt. Ltd.

B-82, G.T.K. Road Industrial Area Delhi - 110 033 (INDIA)

Ph.: 27113030, 45201100, Fax: 27413437, 45201198

WEBSITE : www.rapiddiagnostic.in



## TECHNICAL SERVICE / INSTALLATION REPORT

Customer's Name and Address: Hind labs, Gurugram  
Shri Mata Sheela Devi Swarna Jayanti  
Diagnostic center.

Date: 27/07/2021      TSR No.: **2550**

Inst. Model: URI 600

Cont. Person: Dr. VIJAY K. SHARMA

Inst. S. No.: N1900500410348SE

Contact No.: 8800356516      Email: \_\_\_\_\_

Service Status: Warranty  R/R  A.M.C.  C.M.C.  Chargeable  Others

Call Status: **INSTALLATION**  Breakdown  PM  Routine Service  Other

Problem Reported/ Installation Details:

\* Calibration Problem,  
Calibration Not okay

Instrument attended and is working satisfactorily

Job to be continued

Following parts are required to be replaced

Following parts have been replaced

Observations: \* Dirty lens

\* Calibration test okay  
after maintenance

Sr. No.	Parts	Old Hologram	New Hologram	Price (Rs.)
1.				
2.				
3.				
4.				
5.				
Total (Rs.)				

Action Taken: \* Cleaned the dirty lens

\* Ran Calibration Strip  
thrice

Customer's Remarks (If any): \_\_\_\_\_

Engineer's Remarks (If any): Machine Working  
properly

Engineer's Name: \_\_\_\_\_

No Payment Done - Warranty/Rental

Engineer's Signature: \_\_\_\_\_

Customer's Signature & Seal \_\_\_\_\_

Sunit  
27/7/21

5

Date:27-07-2021 11:25  
Calibration OK

Date:27-07-2021 11:27  
Calibration OK

Date:27-07-2021 11:29  
Calibration OK

*Sund*  
*27/7/21*

Normal IQC.

Date: 06-07-2022 15:47  
No. : 1101  
ID: -----  
UBG Normal 3.4umol/L  
BIL Neg  
KET Neg  
BLD Neg  
PRO Neg  
NIT Neg  
LEU Neg  
GLU Neg  
SG <=1.005  
pH <=5.0

HDFC BANK  
D 03/2022  
HDFC BANK  
D 03/2022  
HDFC BANK  
D 03/2022

Soniya  
6-7-22

Mamish  
6-7-22

High IQC.

Date:	06-07-2022 15:48
No. :	1102
ID:	-----
UBG	Normal 3.4umol/L
BIL	Neg
KET	Neg
*BLD	2+ Ca80 Ery/uL
PRO	Neg
*NIT	Pos
LEU	Neg
GLU	Neg
SG	1.010
pH	<=5.0

SANK D 03/2022 HDPC BANK D 03/2022

Sonja  
6-7-22

Manish  
6-7-22



Normal

Date:07-07-2022 11:13	D 03/2022
No. :1109	HDFC BANK
ID:-----	D 03/2022
UBG Normal 3.4umol/L	HDFC BANK
BIL Neg	D 03/2022
KET Neg	HDFC BANK
*BLD 2+ Ca80 Ery/uL	D 03/2022
PRO Neg	HDFC BANK
*NIT Pos	D 03/2022
*LEU +- Ca15 Leu/uL	HDFC BANK
GLU Neg	D 03/2022
SG <=1.005	
pH <=5.0	

High IgG.

Raja  
7-7-22

Mamta  
7-7-22

Date:07-07-2022 11:12  
No. :1108  
ID:-----  
UBG Normal 3.4umol/L  
BIL Neg  
KET Neg  
BLD Neg  
PRO Neg  
NIT Neg  
LEU Neg  
GLU Neg  
SG <=1.005  
pH <=5.0

D 03/2022  
HDFC BANK  
D 03/2022  
HDFC BANK  
D 03/2022

Normal IQC.

Zia  
7-7-22

Mamishy  
7-7-22