

## CALIBRATION CERTIFICATE

To,

This is to Certify that Biochemistry Analyzer Meril **CliniQuant Micro** Instrument Sr.No. **CM0323124** Installed at **Raj Lab Services** has been calibrated by our engineer on **09-Jul-2022** with appropriate and sophisticated tools. Bellow parameters are been calibrated:

### TEST ITEMS AND RESULT

Appearance	OK
AD Value Range (120<AD<350)	OK
Temperature Checking (at 37'C)	OK
Printer Test	OK
Aspiration Volume Range	OK
Stability of Pot. Dichromate	OK
Absorbency Linearity at 340	OK
PC Communication Checking	OK
Sample Test 1	OK
Sample Test 2	OK
Sample Test 3	OK

This Calibration Certificate is Valid Up to **08-Jul-2023** .

Conclusion: All parameter are within range. The Machine is working properly. (As per Meril terms.)

Engineer Name: **Mr. Vishnu Gopalakrishnan**

Engineer Sign:

Stamp:



T +91 260 3052100  
F +91 260 3052125

[www.merillife.com](http://www.merillife.com)

Meril Diagnostics Pvt. Ltd.  
Survey No.135/2/B, Bilakhia House  
Muktanand Marg, Chala  
Vapi - 396191. Gujarat, India.