

07-06-2022 TO 07-06-2023

To

Srisamraj Health Services PVT LTD
2nd Floor, 217, salai street
Ramanathapuram-623501


CALIBRATION CERTIFICATE

Instrument : Electrolyte Analyzer with Hco3
Model : ST200 AQUB BICARBONATE
Serial No : STAQUB-145
Make : SensaCore, India

This is certify that the ST200 AQUB BICARBONATE has undergoes Automatic Calibration regularly and hence manual calibration is not required. Instrument is working with the reliable accuracy limits.

Thanking you,

Yours Truly,


Kuppuswamy
Sr. Service Engineer
For DiaReK Health Care Pvt. Ltd.
DiaReK Healthcare (P) Ltd.,
Old No. 11, New No. 13,
Nadar Street,
Behind City Centre,
Mylapore, Chennai- 600 004.
Ph: 24990489, 24987084, 8939656201

DiaReK Health Care Pvt. Ltd.

PRINTER : ON

Jul-09-22 16:04:08

QC test : 108

Date time: 2022-06-15 13:05:09

QC Level : QC Level 3

Lot Number :

Operator ID: saranya

TRef : 27.3

ELECTROLYTES

Na = 171.9 mmol/L

K = 6.00 mmol/L

iCa = 0.77 mmol/L

pH = 7.937

HCO3 = 12.50 mmol/L

Cl = 126.8 mmol/L

Li = NA mmol/L

CALCULATED PARAMETERS

pCO2 = 5.81 mmHg

TCO2 = 12.68 mmol/L

BE = 0.00 mmol/L

Anion Gap

AG-Na = 32.66 mmol/L

AG-K = 38.66 mmol/L