



L-95, 5th Cross, 1st Main, Kirloskar Colony 3rd Stage,
Water Tank Road, Basaveshwaranagar, Bangalore-560079.
+91 080-2322 3936, 96633 04352
calibration@sarvashree.com
www.sarvashree.com



Cert No. : CC-2291

NABL Accredited Calibration Services as per ISO/IEC 17025:2017

CALIBRATION CERTIFICATE

SSI/FF-20/v1

Page 1 of 2

1 Name and Address of Customer : M/s. Mediquest Healthcare.,
Mahim Complex, Pumpwell Old Road,
Near Cochin Bakery, Kankanady,
Mangalore 575002.

2 Customer Reference

2.1 SRF No : 6223 Dated : 25 June 2022
2.2 Certification No. : SS/22/6223-07 ULR No. : CC2291220000007882F
2.3 Date of Calibration : 27 June 2022
2.4 Next Calibration Due : 26 June 2023

3 Details Of Device Under Calibration(DUC).

3.1 Nomenclature : Variable Micropipette Model : Finnpiette F3
3.2 Make : Thermo Scientific ID No: ----
3.3 SI No : RW12196 Range : 100-1000 µl
3.4 No. of Pages : 2 LC: 0.2 µl
3.5 Calibration Procedure No. : SOP-MECH-03
3.6 DUC Condition : Satisfactory
3.7 Calibration done at : Mech Lab, Sarvashree

4 Environmental Condition

Temperature 20 ± 1.5 °C Humidity 40 to 60 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No	Traceable Cert. No.	Validity
1	Electronic Balance	Radwag- AS82/220.R2	585650	TVCSPL 22/03/070-02	15-Mar-23

6 Note:

- 6.1. Measurement uncertainty reported is at approx 95% confidence level with coverage factor k=2.
6.2. Kindly refer to Note (s) section mentioned over leaf.


Calibrated By


Rakesh J

(Calibration Engineer)



Authorised By


Noushad N

(Lab In-charge)