



L-95, 5th Cross, 1st Main, Kirloskar Colony 3rd Stage,
Water Tank Road, Basaveshwaranagar, Bangalore-560079.

+91 080-2322 3936, 96633 04352

callbration@sarvashree.com

www.sarvashree.com



Cert No. : CC-2291

NABL Accredited Calibration Services as per ISO/IEC 17025:2017

CALIBRATION CERTIFICATE

SS/FF-20/v1

Page 1 of 2

Name and Address of Customer : M/s. MEDIQUEST HEALTHCARE.,
Mahim Complex, Pumpwell Old Road,
Near Cochin Bakery, Kankanady, Mangalore -575002.

Customer Reference

2.1 SRF No : 6220
2.2 Certification No. : SS/22/6220-01
2.3 Date of Calibration : 25 June 2022
2.4 Next Calibration Due : 24 June 2023

Dated : 25 June 2022
ULR No. : CC229122000007886F

Details Of Unit Under Calibration(UUC).

3.1 Nomenclature : Centrifuge
3.2 Make : Remi Model : R-8C-Plus
3.3 SI No : ZGBN-05284 ID. No. : --
3.4 No. of Pages : 2 Location: --
3.5 Calibration Procedure No. : SOP-MS-07
3.6 DUC Condition : Satisfactory
3.7 Calibration done at : Onsite

Environmental Condition

4.1 Temperature : 25 ± 5 °C Humidity : 40 to 60 %Rh

Standards Used for calibration

| Sl. No. | Nomenclature | Make & Model | SI. No | Traceable Cert. No. | Validity |
|---------|--------------|--------------|---------|---------------------|-----------|
| 1 | Tachometer | Lutron | Q652908 | TVCSPL 22/01/155-01 | 30-Jan-23 |

standard followed -SANAS TR 45-01

Note:

- 6.1. Measurement uncertainty reported is at approx 95% confidence level with coverage factor k=2
- 6.2. Kindly refer Note(s) Section mentioned over leaf.

Calibrated By

Rakesh J

(Calibration Engineer)



Authorised By

Noushad N
(Lab In-Charge)

NOTE:

1. Measurement uncertainty reported is at approx 95% confidence level with coverage factor $k=2$
2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of Sarvashree.
3. The Calibration Certificate relates only to the above DUC. DUC Indicates Device Under Calibration.
4. Corrections/Erasing invalidate the calibration certificate.
5. All Standards / Masters used for calibration are traceable to National / International Standards.
6. Any error in this cert should be brought to our knowledge within 45 days from the date of this certificate.
7. Results reported are valid at the time of and under stated conditions of measurements.
8. Conformity statement is given only when requested by the customer.
9. NABL-133 Guidelines are adopted for use of NABL Symbol.

CAL CERT. NO. SS/22/6220-01

ULR No : CC229122000007886F

Page No: 2 of 2

Results :

Range : 6000 RPM


L C 10 RPM

| SI No. | DUC Set | DUC Reading | Standard Readings | Deviation Observed | Expanded Uncertainty Uo |
|--------|---------|-------------|-------------------|--------------------|-------------------------|
| | in RPM | in RPM | in RPM | in RPM | in \pm RPM% |
| 1 | 500 | 500 | 500.5 | -0.5 | 0.90 |
| 2 | 1000 | 1000 | 1002 | -2 | |
| 3 | 1500 | 1500 | 1503 | -3 | |
| 4 | 3000 | 3000 | 3004 | -4 | |
| 5 | 5000 | 5000 | 5007 | -7 | |
| 5 | 6000 | 6000 | 6008 | -8 | |

Conclusion Remarks


- 1) Calibration done without any adjustments.
- 2) DUC indicates Device Under Calibration

Calibrated By


Rakesh J
(Calibration Engineer)



Authorised By


Noushad N
(Lab In-Charge)