



L-95, 5th Cross, 1st Main, Kirloskar Colony 3rd Stage,  
Water Tank Road, Basaveshwaranagar, Bangalore-560079.  
+91 080-2322 3936, 96633 04352  
calibration@sarvashree.com  
www.sarvashree.com



Cert No. : CC-2291

NABL Accredited Calibration Services as per ISO/IEC 17025:2017

CALIBRATION CERTIFICATE

SS/FF-20/v1

Page 1 of 2

**1 Name and Address of Customer :** M/s. MEDIQUEST HEALTHCARE.,  
Mahim Complex, Pumpwell Old Road,  
Near Cochin Bakery, Kankanady, Mangalore -575002.

**2 Customer Reference**

2.1 SRF No : 6223 Dated : 27 June 2022  
2.2 Certification No. : SS/22/6223-01 ULR. No : CC229122000007876F  
2.3 Date of Calibration : 27 June 2022  
2.4 Next Calibration Due : 26 June 2023

**3 Details Of Device Under Calibration(DUC).**

3.1 Nomenclature : Timer Model : --  
3.2 Make : -- ID No. --  
3.3 SI.No : -- Location: --  
3.4 No.of Pages : 2  
3.5 Calibration Procedure No. : SOP-ETH-28  
3.6 DUC Condition : Satisfactory  
3.7 Calibration done at : lab, Sarvashree

**4 Environmental Condition**

4.1 Temperature 25 ± 4 °C Humidity 30 to 75 %Rh

**5 Standards Used for calibration**

Sl. No.	Nomenclature	Make & Model	SI. No	Traceable Cert. No.	Validity
1	Digital Timer	Racer / --	---	CC2101023/1	2-Jul-22

**6 Note:**


- 6.1. Measurement uncertainty reported is at approx 95% confidence level with coverage factor K=2
- 6.2. Kindly refer Note(s) Section mentioned as below.

Calibrated By

  
Rakesh J  
(Calibration Engineer)



Authorised By

  
Sanath K S  
(Lab In-Charge)

CAL CERT. NO. SS/22/6223-01 ULR. No: CC229122000007876F

Page No: 2 of 2

Range : 0 to 60 min

L.C. : 1 sec

Results :

Parameter	DUC Reading	Standard Reading	Observed Deviation	Expanded Measurement Uncertainty $\pm$ %
Digital Timer	5 sec	5.00 sec	0.00 sec	0.56
	10 sec	10.00 sec	0.00 sec	0.56
	15 sec	15.00 sec	0.00 sec	0.56
	30 sec	30.01 sec	-0.01 sec	0.56
	60 sec	60.02 sec	-0.02 sec	0.56

Conclusion Remarks :


- 1) Calibration done without any adjustments.
- 2) DUC indicates Device Under Calibration

Calibrated By

  
Rakesh J  
(Calibration Engineer)



Authorised By

  
Sanath K S  
(Lab In-Charge)





L-95, 5th Cross, 1st Main, Kirloskar Colony 3rd Stage,  
Water Tank Road, Basaveshwaranagar, Bangalore-560079.  
+91 080-2322 3936, 96633 04352  
calibration@sarvashree.com  
www.sarvashree.com



Cert No. : CC-2291

NABL Accredited Calibration Services as per ISO/IEC 17025:2017

CALIBRATION CERTIFICATE

1 Name and Address of Customer : M/s. MEDIQUEST HEALTHCARE.,  
Mahim Complex, Pumpwell Old Road,  
Near Cochin Bakery, Kankanady, Mangalore -575002.

2 Customer Reference  
2.1 SRF No : 6223 Dated : 27 June 2022  
2.2 Certificate No. : SS/22/6223-08 ULR No : CC229122000007883F  
2.3 Date of Calibration : 27 June 2022  
2.4 Next Calibration Due : 26 June 2023

3 Details Of Device Under Calibration(DUC),  
3.1 Nomenclature : Timer Model : 105  
3.2 Make : Winner Loudstar ID. No. : --  
3.3 Sl.No : -- Range : 0 to 99.99 mins  
3.4 No of Pages : 1 LC : 1 s  
3.5 Calibration Procedure No. : SOP-ETH-28 Location : --  
3.6 DUC Condition : Satisfactory  
3.7 Calibration done at : Lab, Sarvashree

4 Environmental Condition  
4.1 Temperature 25 ± 10 °C Humidity 30 to 75 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make / Model	Sl. No	Traceable Cert. No.	Validity
1	Digital Stop Watch	Racer/HS-80TW	--	CC2101023/1	2-Jul-22

Calibration Results :-

SL No.	DUC Reading in H:min:sec	Standard Reading in H:min:sec	Observed Deviation H:min:sec	Measurement Uncertainty ±
1	00:05:00	00:05:00	00:00:00	0.56%
2	00:10:00	00:10:01	00:00:01	
3	00:15:00	00:15:01	00:00:01	
4	00:30:00	00:30:05	00:00:05	
5	00:60:00	00:59:88	00:00:12	

6 Note:

- 6.1. Measurement uncertainty reported is at approx 95% confidence level with coverage factor K=2
- 6.2. Kindly refer Note(s) Section mentioned as below.

Calibrated By  
  
Rakesh J  
(Calibration Engineer)



Authorised By  
  
Sanath K S  
(Lab In-Charge)