

Ph. : (O) + 91 265 3927448/49/50
2632148/49/50
Fax : (O) + 91 265 2632151

LABLINE EQUIPMENTS PVT. LTD.

An ISO 9001 : 2008 & 13485: 2003 Accredited Company

Mfgr. Of Medical Equipments

614/1, G.I.D.C. B/H Himalaya Machinery, Makarpura, Vadodara - 390 010, Gujarat, INDIA
E-mail : haresh@lablineindia.com • mail@lablineindia.com • Web Site: www.lablineindia.com



CALIBRATION CERTIFICATE

CUSTOMER :

CERTIFICATE NO. : **LL/006/22-23**

PATHOHUB BIOMEDICAL LLP
G-43, SUMEL BUSINESS PARK 6,
SHAHIBAUG,
AHMEDABAD - 380004

DESCRIPTION OF ITEM : Brushless Medical Centrifuge SC 888 With Rotor RT 08 (24 x 5 ml Swing Out Rotor)

CALIBRATION DATE : 29/08/2021

MODEL NO.: SC 888 M/C.Sr.No. 647 A RANGE : 500 - 3500 RPM

DETAILS OF EQUIPMENTS USED FOR CALIBRATION :--

SR. NO.	DESCRIPTION	CERTIFIED BY	CERTIFICATE NO.	VALIDITY
1	DIGITAL TACHOMETER	G.I.I.	I - 3776	1 YEAR

OBSERVATION :

R.P.M. CALIBRATION * [For Centrifuges only]

POINT	INSTRUMENT R.P.M. DISPLAYED	CALIBRATED TACHOMETER READING	ERROR OBSERVED	ACCEPTED ERROR	RESULT
1	500	495	5	+/- 100	OK
2	2500	2510	-10	+/- 100	OK
3	3500	3490	10	+/- 100	OK

TEMPERATURE CALIBRATION * : [For Dry bath incubators only]

POINT	INSTRUMENT TEMPERATURE DISPLAYED (°C)	CALIBRATED THERMOMETER TEMP. (°C)	ERROR OBSERVED	ACCEPTED ERROR	RESULT

NEXT CALIBRATION DUE ** :

APPROVED BY :

Labline Equipments Pvt. Ltd.

Director
Labline Equipments Pvt. Ltd
614/1, G.I.D.C. Estate,
Makarpura, Vadodara - 390 010
Phone: 3927448, 49, 50.

INSPECTION DEPARTMENT

* : RPM & temperature readings are taken on "LABLINE EQUIPMENTS PVT. LTD." standard testing panels and instruments. All readings are taken under ideal test condition.

** : As required by the user.