



Prism Calibration Centre

www.primscalibration.com, www.eindustries.in

E-mail : info@primscalibration.com, info@eindustries.in



NABL Accredited Calibration Lab

Page No.: 1 of 2

CALIBRATION CERTIFICATE

Calibrated for (Name & Address) Shraddhadeep Green Cross Pathology Laboratory 604-607, A Wing, 6th Floor, Vrundavan Trade Centre, Nr. Reliance Cross Road, Kudasán Gandhinagar Gujrat India	ULR Number : CC248021000006161F Issue Date : 25/05/2021 Certificate No. : 20210524-1380-2 Item Received Date. : 24/05/2021 Calibration Date. : 24/05/2021 Next Due on. : 23/05/2022 Calibration Location : Lab
Condition of Item On received : Satisfactory	

DETAILS OF UNIT UNDER CALIBRATION (UUC):

Name : Thermo Hygrometer with Sensor	Job No. : 20210524-1380
Make : ACETEQU	Identification No. : THYGS-01
Model : NA	Range : In -10 to 50 °C to Out -40 to 70 °C
Serial No. : NA	Resolution/Least Count : 0.1 °C / 1 %RH
Type : Digital	Accuracy(±) : NA
Location :	

Calibration Procedure Used : PRISM/CAL/SOP/THG/10 Calibration IS Standard Used :

Environmental Conditions:
Ambient Temp: 24.5 °C Humidity: 53.8 %RH

DETAILS OF STANDARD USED FOR CALIBRATION

NAME	ID. NO.	CERTIFICATE NO.	CALIBRATED BY	VALID UPTO
Digital Hygrometer	PRISM/TEMP/HYG/234	CC210120000002473F	DVG Laboratories & Consultants Pvt. Ltd.	09/11/2021

CALIBRATION RESULTS:

In Temperature Readings@50%RH | Discipline:Group :- Thermal : Temperature

Sr. No.	Ser Point On Humidity Chamber	Average Reading On Master	Average Reading On UUC	Deviation in Reading	Expanded Uncertainty
	°C	°C	°C	°C	°C
1	25.0	25.12	25.8	0.68	±0.77
2	35.0	35.21	35.9	0.69	±0.77
3	40.0	39.88	40.8	0.92	±0.77


Out Temperature Readings@50%RH | Discipline:Group :- Thermal : Temperature

Sr. No.	Ser Point On Humidity Chamber	Average Reading On Master	Average Reading On UUC	Deviation in Reading	Expanded Uncertainty
	°C	°C	°C	°C	°C
1	25.0	25.14	25.6	0.46	±0.77
2	35.0	35.18	35.8	0.62	±0.77
3	40.0	39.88	40.5	0.62	±0.77

PRISM/TEMP/THG/01

Calibrated By : Rushabh Gupta
Calibration Engineer

Authorised By : D.L. Panchal
Lab Incharge



Inst. Supply & Calibration, Validation, GPCB Approved Env. Auditor

F/101, Rudraksh Complex-II, Beside Syndicate Bank, Nr. Jasoda Nagar Cross Road-III, GIDC, Vatva, A'bad-382 445.

Customer Care : 9099062851, 9099062852, Tech. Help Desk 7878991188

Prism Calibration Centre

www.prismcalibration.com, www.eindustries.in

E-mail : info@prismcalibration.com, info@eindustries.in



NABL Accredited Calibration Lab CALIBRATION CERTIFICATE

Certificate No.:20210524-1380-2 Calibration Date:24/05/2021 Next Due on:23/05/2022 Page No.:2 of 2

Humidity Readings@25°C Discipline:Group :- Thermal : Humidity					
Sr. No.	Ser Point On Humidity Chamber	Average Reading On Master	Average Reading On UUC	Deviation in Reading	Expanded Uncertainty
	% RH	% RH	% RH	% RH	% RH
1	35	35.22	33	-2.22	±1.7
2	50	50.35	47	-3.35	±1.7
3	75	74.88	72	-2.88	±1.7

Remarks


- *These Results Are Obtained At The Time Of Calibration Only Relate to the item Calibrated.
 - *The Calibration Certificate Shall not be reproduced except in full without written approval of Prism Calibration Centre.
 - * UUC = Unit Under Calibration. Above Instrument Calibrated at Lab.
 - * All the measurement is traceable to national & international Standards.
 - * Calibration Procedure No.For Lab:- PRISM/CAL/SOP/THG/10,As per Comparison method NABL-141 .
 - *The Value Have been rounded off as per IS:2-1960 wherever applicable.
- The reported expanded uncertainty of measurement is calculated at approximately 95.45% of confidence level with coverage factor $k = 2$.

PRISM/TEMP/THG/C1

Calibrated By : Rushabh Gupta
Calibration Engineer

Authorised By: D L Panchal
Lab Incharge

----- End of Certificate -----



Inst. Supply & Calibration, Validation, GPCB Approved Env. Auditor

F/101, Rudraksh Complex-II, Beside Syndicate Bank, Nr. Jasoda Nagar Cross Road-III, GIDC, Vatva, A'bad-382 445.

Customer Care : 9099062851, 9099062852, Tech. Help Desk 7878991188



Prism Calibration Centre

www.prismcalibration.com, www.eindustries.in

E-mail : info@prismcalibration.com, info@eindustries.in



NABL Accredited Calibration Lab

Page No.: 1 of 2

CALIBRATION CERTIFICATE

Calibrated for (Name & Address) Shradhdheep Green Cross Pathology Laboratory 604-607, A Wing, 6th Floor, Vrundavan Trade Centre, Nr. Reliance Cross Road, Kudasán Gandhinagar Gujrat India .	ULR Number : CC248021000006160F
	Issue Date : 24/05/2021 Certificate No. : 20210524-1380-1 Item Received Date. : 24/05/2021 Calibration Date. : 24/05/2021 Next Due on. : 23/05/2022 Calibration Location : Lab
Condition of Item On received : Satisfactory	

DETAILS OF UNIT UNDER CALIBRATION (UUC):

Name : Thermo Hygrometer	Job No. : 20210524-1380
Make : BEETECH	Identification No. : THYG-01
Model : NA	Range : -50°C TO 70°C to 10 TO 99 %RH
Serial No. : NA	Resolution/Least Count : 0.1 °C / 1 %RH
Type : Digital	Accuracy(±) : NA
Location :	

Calibration Procedure Used : PRISM/CAL/SOP/THG/10 Calibration IS Standard Used .:

Environmental Conditions:
 Ambient Temp: 24.5 °C Humidity: 54.5 %RH

DETAILS OF STANDARD USED FOR CALIBRATION

NAME	ID. NO.	CERTIFICATE NO.	CALIBRATED BY	VALID UPTO
Digital Hygrometer	PRISM/TEMP/HYG/234	CC210120000002473F	DVG Laboratories & Consultants Pvt. Ltd.	09/11/2021

CALIBRATION RESULTS:

Temperature Readings@50%RH | Discipline:Group :- Thermal : Temperature

Sr. No.	Ser Point On Humidity Chamber	Average Reading On Master	Average Reading On UUC	Deviation in Reading	Expanded Uncertainty
	°C	°C	°C	°C	°C
1	25.0	25.12	24.9	-0.22	±0.77
2	35.0	35.21	35.4	0.19	±0.77
3	40.0	39.88	40.2	0.32	±0.77

Humidity Readings@25°C | Discipline:Group :- Thermal : Humidity


Sr. No.	Ser Point On Humidity Chamber	Average Reading On Master	Average Reading On UUC	Deviation in Reading	Expanded Uncertainty
	% RH	% RH	% RH	% RH	% RH
1	35	35.23	36	0.77	±1.7
2	50	50.38	53	2.62	±1.7
3	75	74.89	78	3.11	±1.7

Remarks

PRISM/TEMP/THG/01

Calibrated By : Rushabh Gupta
 Calibration Engineer

Authorized By : D L Panchal
 Lab-Incharge



Inst. Supply & Calibration, Validation, GPCB Approved Env. Auditor

F/101, Rudraksh Complex-II, Beside Syndicate Bank, Nr. Jasoda Nagar Cross Road-III, GIDC, Vatva, A/bad-382 445.

Customer Care : 9099062851, 9099062852, Tech. Help Desk 7878991188