



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC23392200003271F	Calibration Done on : 07.07.2022	S.R.F.No.	0001/07
Issue Date : 08.07.2022	Calibration Valid Upto : 06.07.2023	S.R.F Date	: 07.07.2022
UUC Receipt Date: 06.07.2022	Certificate No. : PCS/3271	Page	: 01 of 01
Calibrated for M/S. DR. BATRA'S LABORATORY 102, SAMRAT CHAMBERS,OPP. MADHYAVARTI SCHOOL, POONAM STUDIO LANE,DANDIA BAZAR,VADODARA -390001			

Identification on UUC (Unit Under Calibration)

Instrument Name	Centrifuge	Size/Range	0 to 6000 RPM
Least Count	100 RPM	Conition Of Item	Good
Manufactured By	Remi	Calibration Performed at	Lab
Manufacturer M/Sr. No.	R-4C / DCLC-263	Location	Lab
Party ID Mark	CF - 01	Ref. Standard/Procedure	IS: 12508

Calibration Procedure / Method : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-24

Environmental conditions (s) : Temperature : 24 ± 3 °C . Relative Humidity : 50 +/-10 % RH

Standard & Major Equipment(s)

Standard(s)& Equipments	Range / Serial No.	Calibrating Agency	Traceability Cert. No.	Certificate Date	Valid Up to
Dig. Tachometer	Upto 99999.9/ 109-0595Y	Transcal	CC22312200003 2672F	19.03.2022	18.03.2023

Calibration Results:

U.U.C Value (RPM)	Standard Value (RPM)	Error (RPM)	Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k=2) = ±
1000	1001.6	-1.6	5.67%
2000	2003.8	-3.8	2.84%
3000	3005.4	-5.4	1.90%
4000	4007.6	-7.6	1.42%
5000	5009.1	-9.1	1.14%

***** END OF REPORT *****

NOTE:

- The result are valid only for this Calibrated item.
- The calibration report shall not be reproduced without the written approval of calibration laboratory.
- Result reported are valid at the time of & under the stated conditions of measurement
- Laboratory standards are traceable to National/International Standards.
- Calibration Certificates issued for Weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scales / Tapes etc. are for scientific purpose only and should not be used for Trade / Commercial use.

AUTHORISED SIGNATORY

[Signature]

M.A. BIDDICH
(CEO)

CALIBRATED/CHECKED BY

[Signature]

