



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233922000003268F	Calibration Done on : 07.07.2022	S.R.F.No. : 0001/07
Issue Date : 08.07.2022	Calibration Valid Upto : 06.07.2023	S.R.F Date : 07.07.2022
UUC Receipt Date: 06.07.2022	Certificate No. : PCS/3268	Page : 01 of 01
Calibrated for M/S. DR. BATRA'S LABORATORY 102, SAMRAT CHAMBERS,OPP. MADHYAVARTI SCHOOL, POONAM STUDIO LANE,DANDIA BAZAR,VADODARA -390001		

Identification on UUC (Unit Under Calibration)

Instrument Name	Micro Pipette	Size/Range	1000 µl
Least Count	N/A	Condition of Item	Good
Manufactured By	---	Calibration Performed at	Lab
Manufacturer M/Sr.No.	---	Location	Lab
Party ID Mark	MP - 03	Ref. Standard	ISO-8655-6

Calibration Procedure / Method : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-51

Environmental conditions (s) : Temperature: 23 ± 3 °C. Relative Humidity : 50 ± 10% RH

Standard & Major Equipment(s) Used :

Standard Equipment	Range/ Serial No.	Calibrating Agency	Traceability Cert. No.	Certificate Date	Valid Up to
Standard Weight	1 mg to 200 g	Weightronics	WMCL/E/2020-06/3369	03.06.2020	02.06.2023

Calibration Results at 27 °C :

Nominal Value UUC (µl)	Observed Value (µl)	Error (µl)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ±
1000	1001.96	-1.96	0.38 µl

***** END OF REPORT *****

NOTE:

- The result are valid only for this Calibrated item.
- The calibration report shall not be reproduced without the written approval of calibration laboratory.
- Result reported are valid at the time of & under the stated conditions of measurement
- Laboratory standards are traceable to National/International Standards.
- Calibration Certificates issued for Weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scales / Tapes etc. are for scientific purpose only and should not be used for Trade / Commercial use.

AUTHORISED SIGNATORY

CALIBRATED/CHECKED BY

