



C-2 Shriram Complex, Nyay Khand-I, Indirapuram, Ghaziabad-201014 (U.P) INDIA Ph. No. +91-9899380410, +91-9716668085, +91-7669021291-95

Calibration Certificate

Page No.: 01 of 01

Format No.: HY/FM/33		
Certificate No.	HY/21/1371-151	CALIBRATION CERTIFICATE OF
Date of Calibration :	26.11.2021	
Next Date for Calibration	26.11.2022	
(As per agreed with the Customer	.)	
Cartificata Issua Data	27 11 2021	Centrifuge Machine

CUSTOMER ADDRESS		DESCRIPTION OF DEVICE UNDER CALIBRATION		
M/s Care Hospital		Make / Model	Remi	
(A Unit of Goel Superspeciality	Medical Centre Pvt.Ltd.)	Instrument Sr. No.	NA	
Near Delapeer Petrol Pump, Stadium Road, Bareilly - 243001 INDIA		Resolution / Deast count	As Per Range	
litear belapeer retror ampyonana	,	Range / Size	As Per Instrument	
		Asset Code / I.D No.	CH/LAB/CF/01	
Service Request / Job No.	1371/1371-151	Location	LAB	
Service Request Dated	24.11.2021	Instrument Condition	In Working	

	11V (MI /2M1 4	Calibration Perfo	ormed At		At Site
Calibration Procedure	HY/WI/2M14	Cambration refre	Cambration reflormed ne		Medical
Reference Standard	NABL-126	Discipline			153 AFT AFT
ENVIRONMENTAL CONDITIONS	Temperature	(25 ± 15) °C	Relative Hum	idity	(60 ± 15) %RH

DETAILS OF REFERENCE STANDARDS AND MAJOR EQUIPMENTS USED FOR CALIBRATION

DETAILS OF REFERENCES TIMES THE				D . D . D . t .	
Sr. No	Instrument Details	Certificate No.	Calibrated By	Calibration Due Date	
1		CC205520000000753F	TNC	18.01.2022	
1 1 1	Digital Tachometer	0020002000			

The Standards used for calibration are traceable to National Standards.

CALIBRATION RESULT

PARAMETER	DUC VALUE (rpm)	STANDARD VALUE (rpm)	ERROR (rpm)	Uncertainty ± (%)
Speed	800	803.7	-3.70	0.70
*	1000	1002	-2.00	0.70
	1500	1505	-5.00	0.70
	2500	2506	-6.00	0.70
	3000	3003	-3.00	0.70

Expande	ed Uncertainty of measurment at approximately 95% Confidence Level with coverage factor $k=2$	
Note :-	1. DUC = Device Under Calibration 2. The Certificate refers only to the particular items submitted for Calibration. 3. This Certificate shall not be reproduced, except in full, without per in the second of the s	(Calib Engineer) APRROVED BY



