

**To,  
Care Hospital  
Near Delapeer Petrol pump,  
Stadium Road, Bareilly (U.P)-243001**

## **CALIBRATION CERTIFICATE**

This is to certify that Electrolyte analyzer ST-200 (Sr.No.-ST-5727) is exclusively marketed in India by PSR Diag. Pvt. Ltd. is installed at care Hospital Bareilly

We are solely responsible for maintenance (preventive and breakdown) and calibrations as recommended by the manufacturers in the operator manual and updates received from them time to time.

Instrument hardware calibration is checked & done by our Instrument Engineer on 14.06.2022 & next calibration is due on 13.12.2022.

**PSR DIAG. Pvt. Ltd.,**



**Authorized Signatory**

## **PSR DIAGNOSTIC PVT. LTD.**

PSR PLUS  
LEVEL 2

---

ANALYZE QC

---

Na = 148.2 mmol/L  
K = 4.25 mmol/L  
iCa = 1.25 mmol/L

---

Jul-16-22 14:08:22

---

ANALYZE TYPE : SERUM/PLASMA  
Jul-16-22 14:10:40

PSR PLUS  
LEVEL 2

ANALYZE QC

Na = 142.9 mmol/L

K = 4.02 mmol/L

iCa = 1.20 mmol/L

Jul-02-22 17:35:49

PSR PLUS  
LEVEL 2

ANALYZE QC

Na = 145.8 mmol/L  
K = 4.07 mmol/L  
iCa = 1.22 mmol/L

Jul-02-22 17:27:50

PSR PLUS  
LEVEL 2

ANALYZE QC

Na = 143.8 mmol/L  
K = 4.07 mmol/L  
iCa = 1.22 mmol/L

Jul-02-22 17:30:29

PSR PLUS  
LEVEL 2

ANALYZE QC

Na = 146.3 mmol/L  
K = 4.19 mmol/L  
iCa = 1.28 mmol/L

Jul-02-22 17:33:20

PSR PLUS  
LEVEL 2

ANALYZE QC

Na = 148.0 mmol/L  
K = 4.19 mmol/L  
iCa = 1.28 mmol/L

Jul-02-22 17:39:47

ANALYZE TYPE : SERUM/PLASMA  
Jul-02-22 17:40:57

ANALYZE TYPE: QUALITY CONTROL  
Jun-24-22 15:01:43

PSR PLUS  
LEVEL 2

ANALYZE QC

Na = 144.9 mmol/L  
K = 4.09 mmol/L  
iCa = 1.23 mmol/L

Jun-24-22 15:03:34



ANALYZE TYPE: QUALITY CONTROL

Jun-13-22 15:51:00

-----

PSR PLUS  
LEVEL 2

ANALYZE QC

Na = 147.0 mmol/L

K = 4.11 mmol/L

iCa = 1.28 mmol/L

Jun-13-22 15:53:03

ANALYZE TYPE: QUALITY CONTROL  
May-23-22 16:46:45

---

PSR PLUS  
LEVEL 2

---

ANALYZE QC

---

Na :145.5 mmol/L  
K :4.15 mmol/L  
iCa:1.28 mmol/L

---

May-23-22 16:48:53

---

ANALYZE TYPE : SERUM/PLASMA  
May-23-22 16:49:32

---

ANALYZE TYPE: QUALITY CONTROL  
May-20-22 13:40:23

---

PSR PLUS  
LEVEL 2

---

ANALYZE QC

---

Na :145.0 mmol/L  
K :4.18 mmol/L  
iCa:1.29 mmol/L

---

May-20-22 13:42:06

---