



Case No.	CAS-0014673900	Instrument Model	ELECTROLYTE ANALYZER W/O STARTERKIT 9180
Order No.	ORD-0018944526	Instrument Serial No.	12730
Contract Type	PREVENTATIVE MAINTENANCE & PARTS	Finance Status	CASH
Lab/Inst./hosp.Name	SRL Diagnostic-Dr. Avinash Phadke		
Customer No.	0052086132		
Contact Name :	Thanuja		
Contact Number :	+912225623742		
Address :	Sumati, Ground floor,		
City :	Mulund West		

Call Received Date/Time:	26.12.2021 17:30	Call Attended Date/Time:	09.03.2022 10:00
--------------------------	------------------	--------------------------	------------------

Job Type	PM Visit
Job Description	
Action Summary	Cause: Preventive Maintenance Workdone:1) checked the system & checked all modules, 2) Carried out preventive maintenance satisfactorily as per checklist along with p.m. kit, 3) Performed the calibration & done samples Spare consumable used: Gmmi no. 03074064001 Item: Harness Main Tubing set Qty: 01 Batch: 354121 Note: Replaced under comfort AMC Verification:Machine working ok Customer Satisfaction Rating (1-5):5

Spare Part Replaced

Part No	Parts Description	Batch No	Batch Expiry Date	Qty	Invoice Type
03074064001	HARNESS, MAIN TUBING, 918X			1	Contract

Time Report

Effective Visit Date : 09.03.2022	Complete Date : 09.03.2022
-----------------------------------	----------------------------

Date	Type	Time
2022.03.09	Travel Time--Standard	1
2022.03.09	Working Time--Standard	1
Total:		2

Customer's Signature Name : Thanuja	Service Engineer/Application Specialist Name : Machindar Parappa Bilkar
Date: 09.03.2022	Date: 09.03.2022