



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2022



Lab Name	HARLEY PATH LABS PVT LTD	Lab No	15578
Constituent Group	Chemistry II	Date of Result Entered :	18/06/2022
PT item	Lyophilized Serum	Date of Report Published :	07/07/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	849	86.39	9.86	8.52	98.8 mg/dL	1.46	0.58
2	UREA II	Urease UV / GLDH II	Transasia / Erba	672	44.59	10.74	4.79	51.7 mg/dL	1.48	0.37
3	CREATININE II	Enzymatic Colorimetric II	Transasia / Erba	166	2.02	8.80	0.18	1.94 mg/dL	-0.45	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	735	3.82	10.81	0.41	4.74 mg/dL	2.23	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	723	6.14	8.88	0.55	5.43 g/dL	-1.30	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	700	3.51	8.77	0.31	3.95 g/dL	1.43	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	742	6.84	12.50	0.86	7.7 mg/dL	1.01	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	810	120.81	12.10	14.62	111 mg/dL	-0.67	1.03
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	736	152.58	12.98	19.81	147.2 mg/dL	-0.27	1.46

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
HARLEY PATH LABS PVT LTD
AI, SECTOR-59,
NOIDA
UTTAR PRADESH 201301

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2022



PC-1024

Lab Name **HARLEY PATH LABS PVT LTD** Lab No **15578**
Constituent Group **Chemistry II** Date of Result Entered : **18/05/2022**
PT item **Lyophilized Serum** Date of Report Published : **08/06/2022**

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	826	153.23	10.59	16.23	164.7 mg/dL	0.71	1.13
2	UREA II	Urease UV / GLDH II	Transasia / Erba	628	121.30	12.77	15.49	137.1 mg/dL	1.02	1.24
3	CREATININE II	Enzymatic Colorimetric II	Transasia / Erba	160	9.22	9.43	0.87	8.42 mg/dL	-0.92	0.14
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendressik II	Transasia / Erba	669	1.71	18.52	0.32	1.78 mg/dL	0.22	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	672	6.24	8.05	0.50	5.79 g/dL	-0.90	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	678	3.50	8.73	0.31	3.93 g/dL	1.41	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	721	5.78	13.22	0.76	5.7 mg/dL	-0.10	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	748	121.76	10.15	12.36	116 mg/dL	-0.47	0.90
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	661	89.83	18.41	16.54	69.4 mg/dL	-1.24	1.29

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
HARLEY PATH LABS PVT LTD
A1, SECTOR-59,
NOIDA
UTTAR PRADESH 201301

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2022



PC-1024

Lab Name	HARLEY PATH LABS PVT LTD	Lab No	15578
Constituent Group	Chemistry II	Date of Result Entered :	19/04/2022
PT item	Lyophilized Serum	Date of Report Published :	05/05/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	799	223.18	10.20	22.77	249.5 mg/dL	1.16	1.61
2	UREA II	Urease UV / GLDH II	Transasia / Erba	620	38.07	11.80	4.49	45.4 mg/dL	1.63	0.36
3	CREATININE II	Enzymatic Colorimetric II	Transasia / Erba	154	5.24	8.83	0.46	4.79 mg/dL	-0.97	0.07
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	703	2.90	11.57	0.34	3.49 mg/dL	1.76	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	672	5.80	8.17	0.47	5.77 g/dL	-0.06	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	649	3.38	8.14	0.28	3.48 g/dL	0.36	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	730	7.12	12.13	0.86	7.5 mg/dL	0.44	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	754	111.29	11.48	12.78	107 mg/dL	-0.34	0.93
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	726	114.85	16.13	18.52	94.1 mg/dL	-1.12	1.37

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
HARLEY PATH LABS PVT LTD
AI, SECTOR-59,
NOIDA
UTTAR PRADESH 201301

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2022



Lab Name HARLEY PATH LABS PVT LTD
Constituent Group Chemistry II
PT item Lyophilized Serum

Lab No 15578
Date of Result Entered : 19/03/2022
Date of Report Published : 06/04/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	791	109.92	9.69	10.65	120.6 mg/dL	1.00	0.76
2	UREA II	Urease UV / GLDH II	Transasia / Erba	601	80.81	11.24	9.09	82.9 mg/dL	0.23	0.74
3	CREATININE II	Enzymatic Colorimetric II	Transasia / Erba	141	4.25	7.13	0.30	4.23 mg/dL	-0.07	0.05
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	668	1.57	19.28	0.30	1.84 mg/dL	0.89	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	593	5.34	8.62	0.46	5.57 g/dL	0.50	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	633	3.15	9.30	0.29	3.09 g/dL	-0.20	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	637	8.01	11.65	0.93	8.5 mg/dL	0.53	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	700	105.41	11.28	11.89	98 mg/dL	-0.62	0.90
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	679	188.91	11.47	21.66	180.1 mg/dL	-0.41	1.66

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
HARLEY PATH LABS PVT LTD
AI, SECTOR-59,
NOIDA
UTTAR PRADESH 201301

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - FEBRUARY 2022



Lab Name	HARLEY PATH LABS PVT LTD	Lab No	15578
Constituent Group	Chemistry II	Date of Result Entered :	19/02/2022
PT Item	Lyophilized Serum	Date of Report Published :	05/03/2022

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	728	278.22	10.88	30.27	300.3 mg/dL	0.73	2.24
2	UREA II	Urease UV / GLDH II	Transasia / Erba	579	29.12	11.67	3.40	27.3 mg/dL	-0.54	0.28
3	CREATININE II	Enzymatic Colorimetric II	Transasia / Erba	140	1.06	9.59	0.10	1.02 mg/dL	-0.39	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	614	5.06	11.12	0.56	6.04 mg/dL	1.74	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	604	5.81	9.02	0.52	5.77 g/dL	-0.08	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	597	3.40	8.90	0.30	3.16 g/dL	-0.79	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	524	4.45	12.44	0.55	4.1 mg/dL	-0.63	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	678	112.37	11.73	13.18	101 mg/dL	-0.86	1.01
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	661	88.35	21.56	19.05	64.6 mg/dL	-1.25	1.48

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
HARLEY PATH LABS PVT LTD
AI, SECTOR-59,
NOIDA
UTTAR PRADESH 201301

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2022



Lab Name **HARLEY PATH LABS PVT LTD** Lab No **15578**
 Constituent Group **Chemistry II** Date of Result Entered : **22/01/2022**
 PT item **Lyophilized Serum** Date of Report Published : **07/02/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	687	130.44	8.67	11.30	143 mg/dL	1.11	0.86
2	UREA II	Urease UV / GLDH II	Transasia / Erba	537	47.63	9.86	4.70	52.7 mg/dL	1.08	0.41
3	CREATININE II	Enzymatic Colorimetric II	Transasia / Erba	129	1.73	6.98	0.12	1.74 mg/dL	0.08	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	615	2.30	15.28	0.35	2.2 mg/dL	-0.28	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	556	5.38	9.07	0.49	5.21 g/dL	-0.35	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	577	3.15	8.76	0.28	3.04 g/dL	-0.40	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	592	7.17	12.44	0.89	7.6 mg/dL	0.48	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	636	101.83	11.62	11.83	93 mg/dL	-0.75	0.94
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	577	115.89	15.40	17.85	103.3 mg/dL	-0.71	1.49

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
HARLEY PATH LABS PVT LTD
AI, SECTOR-59,
NOIDA
UTTAR PRADESH 201301

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****