

CERTIFICATE OF CALIBRATION,
QUALITY CONTROL AND MAINTENANCE

To,
Sun Laboratory
Villupuram

This is to certify that the instrument **URA Serial No. 601826022IE**
Manufactured by **Medsource Ozone Biomedicals Pvt.Ltd** is under Calibration.
Quality control check up of the instrument has been the responsibility of
Medsource Ozone Biomedicals Pvt.Ltd

Date of Calibration

:25th June 2022

Calibration & quality control checks as recommended by the manufacturers in
operator manual are being performed regularly and satisfactorily.

This calibration certificate is valid up to 24th june 2023

For **Medsource Ozone Biomedicals Pvt. Ltd.**


Reshma UR
Service Coordinator
Mob: +91-96500 64495
GST No. 06AAAC05845J1ZA



Medsource Ozone Biomedicals Pvt.Ltd

Head office :Plot No.109,HSIIDC,Sector-31,Faridabad-121003,(Haryana),INDIA
Registered Office:238,Sant Nagar,East of Kailash,New Delhi-110065,INDIA
Telephone: +91-129-4286600 | E-mail:mail@ozonebio.com |Website: www.ozonebio.com
CIN:U85100DL2003PTC120213

CERTIFICATE OF CALIBRATION,
QUALITY CONTROL AND MAINTENANCE

To,
Sun Laboratory
Villupuram

This is to certify that the instrument **Eurocount tsSerial No. 850083**
Manufactured by **Medsource Ozone Biomedicals Pvt.Ltd** is under Calibration.
Quality control check up of the instrument has been the responsibility of
Medsource Ozone Biomedicals Pvt.Ltd

Date of Calibration

:25th June 2022

Calibration & quality control checks as recommended by the manufacturers in
operator manual are being performed regularly and satisfactorily.

This calibration certificate is valid up to 24th june 2023

For **Medsource Ozone Biomedicals Pvt. Ltd.**


Reshma UR
Service Coordinator
Mob: +91-96500 64495
GST No. 06AAAC05845J1ZA



Medsource Ozone Biomedicals Pvt.Ltd

Head office :Plot No.109,HSIIDC,Sector-31,Faridabad-121003,(Haryana),INDIA
Registered Office:238,Sant Nagar,East of Kailash,New Delhi-110065,INDIA
Telephone: +91-129-4286600 | E-mail:mail@ozonebio.com |Website: www.ozonebio.com
CIN:U85100DL2003PTC120213

7-2022 GLU-B Quality Control

Batch No.:

Average: 114.00

ZSD Range: 112.00~116.00

No.	Date	Result	Remark
1	16	112.19ng/L	

Print Date: 16-7-2022

7-2022 UREA Quality Control

Batch No.:

Average: 29.00

ZSD Range: 27.00~31.00

No.	Date	Result	Remark
1	16	23.50ng/L	L

Print Date: 16-7-2022

7-2022 CREAT Quality Control

Batch No.:

Average: 1.40

ZSD Range: -0.60~3.40

No.	Date	Result	Remark
1	16	1.76ng/dL	

Print Date: 16-7-2022

7-2022 BIL-T Quality Control

Batch No.:

Average: 1.70

ZSD Range: -0.30~3.70

No.	Date	Result	Remark
1	19	1.52ng/dL	

Print Date: 19-7-2022

7-2022 PROT Quality Control

Batch No.:

Average: 4.90

ZSD Range: 2.90~6.90

No.	Date	Result	Remark
1	19	5.40gn/dL	

Print Date: 19-7-2022

7-2022 ALB Quality Control

Batch No.:

Average: 3.20

ZSD Range: 1.20~5.20

No.	Date	Result	Remark
1	16	3.39gn/dL	

Print Date: 16-7-2022

7-2022 URIC Quality Control

Batch No.:

Average: 4.90

ZSD Range: 2.90~6.90

No.	Date	Result	Remark
1	19	5.18ng/dL	

Print Date: 19-7-2022

7-2022 TCHO Quality Control

Batch No.:

Average: 111.00

ZSD Range: 109.00~113.00

No.	Date	Result	Remark
1	16	111.49ng/dL	

Print Date: 16-7-2022

7-2022 TGL Quality Control

Batch No.:

Average: 151.00

ZSD Range: 149.00~153.00

No.	Date	Result	Remark
1	16	147.39ng/dL	L

Print Date: 16-7-2022