

DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME ALL ANALYSER SUMMARY REPORT - JANUARY 2022



Lab Name

REDCLIFFE LIFETECH PVT LTD

Lab No

15554

Constituent Group

t HbA1c

19/01/2022

PT item

Lyophilized Whole blood

Date of Result Entered

Date of Report Published

07/02/2022

	Constituent	Viethod /						Your		
				Participants						
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	106	5.43	8.22	0.45			0.09
	T)		Beckman AU480/680/5800/DXC700AU	24	5.62	9.29	0.52			0.21
			Any Analyser	219	6.33	14.12	0.89	7.9 %	1.76	0.12

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS:

REDCLIFFE LIFETECH PVT LTD 41/2, POCKET NO.2, VIVEK VIHAR, GMS ROAD, DEHRADUN UTTARAKHAND,248001

> Coordinator Contact details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudosa
Dr. Pamela Christudosa

CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

ALL ANALYSER SUMMARY REPORT - FEBRUARY 2022



Lab Name

REDCLIFFE LIFETECH PVT LTD

Lab No

15554

Constituent

Group

HbA1c

Date of Result Entered

11/02/2022

PT item

Lyophilized Whole blood

Date of Report Published

05/03/2022

			Analyzer Name	No of		Partic	cipants	Your	Your one	
		Principle Name	Secretary and Committee					Value		
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	116	9.83	4.54	0.45			0.08
			Beckman AU480/680/5800/DXC700AU	28	10.23	10.01	1.02			0.39
			Any Analyser	232	10.62	17.04	1.81	10 %	-0.34	0.24

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS:

REDCLIFFE LIFETECH PVT LTD 41/2, POCKET NO.2, VIVEK VIHAR, GMS ROAD,

DEHRADUN UTTARAKHAND,248001

> Coordinator Contact details: Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Panela Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******



DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME ALL ANALYSER SUMMARY REPORT - MARCH 2022



Lab Name

REDCLIFFE LIFETECH PVT LTD

Lab No

15554

Constituent

Group

Date of Result Entered

17/03/2022

PT item

Lyophilized Whole blood

HbA1c

Date of Report Published

06/04/2022

SI.No	Constituent Name	Method / Principle Name	Analyzer Name	No of	DV	Partic	cipants	Your		
î	Manie:			Participants	LOW .	CV	SD	Value		U
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	121	7.19	4.55	0.33			0.06
	1		Beckman AU480/680/5800/DXC700AU	28	7.41	8.01	0.59			0.22
			Any Analyser	212	7.85	10.56	0.83	8.2 %	0.42	0.11

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS:

REDCLIFFE LIFETECH PVT LTD 41/2, POCKET NO.2, VIVEK VIHAR, GMS ROAD, DEHRADUN UTTARAKHAND,248001

> Coordinator Contact details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******



DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME ALL ANALYSER SUMMARY REPORT - APRIL 2022



Lab Name

REDCLIFFE LIFETECH PVT LTD

Lab No

15554

Constituent

HbA1c Group

Date of Result Entered

19/04/2022

PT item

Lyophilized Whole blood

Date of Report Published

05/05/2022

	Constituent	Method i	Analyzer Name	No of	DV	Partic		Your		
		Principle Name		Participants	11.779	CV	SD	Value		
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	121	5.99	5.01	0.30			0.05
			Beckman AU480/680/5800/DXC700AU	31	6.17	7.96	0.49			0.18
			Any Analyser	279	6.67	13.32	0.89	7.7 %	1.16	0.11

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

REDCLIFFE LIFETECH PVT LTD 41/2, POCKET NO.2, VIVEK VIHAR, GMS ROAD, DEHRADUN UTTARAKHAND,248001

> Coordinator Contact details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ****** End of Report *****



DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME ALL ANALYSER SUMMARY REPORT - MAY 2022



Lab Name

REDCLIFFE LIFETECH PVT LTD

Lab No

15554

Constituent Group

HbA1c

Date of Result Entered

14/05/2022

PT item

Lyophilized Whole blood

Date of Report Published

08/06/2022

			Arralyzer Name	No of						
		Principle Name						Value		
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	124	6.90	4.55	0.31			0.06
			Beckman AU480/680/5800/DXC700AU	29	7.24	6.15	0.44			0.17
			Any Analyser	272	7.66	12.08	0.92	7.2 %	-0.50	0.11

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS:

REDCLIFFE LIFETECH PVT LTD 41/2, POCKET NO.2, VIVEK VIHAR, GMS ROAD, DEHRADUN UTTARAKHAND,248001

> Coordinator Contact details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudose

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report ******

1



DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME ALL ANALYSER SUMMARY REPORT - JUNE 2022



Lab Name

REDCLIFFE LIFETECH PVT LTD

Lab No

15554

Constituent

HbA1c Group

Date of Result Entered

17/06/2022

PT item

Lyophilized Whole blood

Date of Report Published

07/07/2022

	Sl.No Constituent Marrie	Method /		Noaf						
	rearing	Principle Name	Constitution (Constitution of Constitution of	Participants				Value		
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	129	7.02	4.34	0.30			0.05
			Beckman AU480/680/5800/DXC700AU	31	7.12	4.49	0.32			0.11
			Any Analyser	279	7.61	11.27	0.86	7.7 %	0.10	0.10

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS:

REDCLIFFE LIFETECH PVT LTD 41/2, POCKET NO.2, VIVEK VIHAR, GMS ROAD, DEHRADUN UTTARAKHAND,248001

> Coordinator Contact details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ****** End of Report ******