

Date: 10-10-2021

## CERTIFICATE OF CALIBRATION

**Customer Name:** "Radcliffe Life Diagnostics", Dehradun.

**Model:** Fully Automated Bio Chemistry Analyzer XL640

**Serial No.:** 640352.

**Calibration done date:** 10/10/2021

**Next Calibration due Date:** On or Before 10/10/2022

**Lab In charge:** Mr. Bantu Yadav.

This is to certify that above mentioned product has been verified of calibration By

Checked Lamp Voltage found in range.

Checked gain of Photometer for all wavelengths(Range450 to950). Values obtained under range.

Checked calibration of probe on all location found o.k.

Checked Temperature at Reagent tray found in range (Range 4 to 12 degree)

and reaction tray found in range (range 36.5 to 37.5 degree).

Checked pressure of pressure tank found in range (range 0.8 to 1.2 bar).

Checked Absorbcuvetteblankswithrespecttoallfiltersfoundinrange0.02to

0.2 Abs.

Followed by calibration of all the three types of chemistries (picked up CHOL, CREAT and SGPT) and QC run with factory values of QC

Calibration at site performed by Mr Tara Singh Mehra

(Tara Singh Mehra )

Area Service Manager,