

Date: 27-10-2021  
Effective Date 27-10-2021

## Certificate of Calibration

**Customer Name:** Redcliffe Lifetech Private Ltd

**Model :** Automated Hematology Analyzer Elite 580

**Serial No. :** K11052125001

**Calibration Done Date:** 27-10-2021

**Next Calibration Due Date On or Before:** 27-10-2022

**Lab In-charge: .** Mr. Banttu Yadav

*This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, MCV and PLT) according to the standard procedures provided by Erba Lachema s.r.o, Karasek.*

Calibration at site performed by  
Engineer Name: Tara Singh Mehra  
Designation: ASM  
Transasia Bio-Medicals Ltd  
Location: Dehradun

Encl: 

1. Certificate of Inspection
2. Assay Sheet of Hematology Calibrator (H Cal)
3. Printouts
4. Traceability Document