

TECHNOCAL

ACCREDITATION FOR MECHANICAL, THERMAL & ELECTRO - TECHNICAL CALIBRATION LABORATORY



CC-2550

CALIBRATION CERTIFICATE				
DISCIPLINE - Non Contact -MECHANICAL-Dimension (Basic-Measuring Instrument , Gauges etc)				
1.CLIENT'S ADDRESS :-		CRF No.	:- CAL-2204-003	
SRL Dr Avinash Phadke pvt. Ltd		Page No.	:- 1 of 1	
Ground floor, Sumati, Behind Saraswat Co-op Bank		ULR No	:- CC25502200000981F	
Mulund-west, Mumbai-400080.		Certificate No.	:- SRL/CAL-2204/03-001	
Envionrmental Condition		Date of Received / Mode	:- 06/04/2022 / By Hand	
Ambient Temp.	:- 24.1 °C	Reference Challen no	:- -	
Relative Humidity	:- 65 %RH	Date of Calibration	:- 06/04/2022	
Location of calibration	:- At site	Next Calibration Due On	:- 05/04/2023	
Condition of Item	:- OK	Date of issue certificate	:- 08/04/2022	
2. Description of Item (UUC)		Calibration method No.	:- TECH-M-WI-06	
Name	:- Centrifuge	Range	:- 100 to 4500 RPM	
Id No	:- Muld-03	L.C.	:- 1 RPM	
Make	:- Remi-R8 C	Location	:- Lab	
Type	:- Non-Contact	Serial no.	:- ZGLN-30679	
Accuracy	:- Not Specified			
3.Detail of Equipment used for calibration				
Name	:- Digital Tachometer (Contact & Non Contact Type)			
Make & Model	:- Lutron & DT-2236			
Certificate No.	:- M-211006-1-1			
Certified By	:- Godrej & Boycee Mfg.Co.Ltd			
ID/Sr. No.	:- TECH/MTM/001, AC05002			
Calibration Validity	:- 13/10/2022			
4.Calibration Results :-				
Calibration Points RPM	Standard Reading RPM	UUC Reading RPM	Error in RPM	Expanded Uncertainty in ± RPM
1000	1013	1000	-13	4.39
2000	2014	2000	-14	4.39
3000	3015	3000	-15	4.39
*The readings of RPM represents the mean of five readings.				
Note:				
1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95% for normal distribution				
2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.				
3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Technocal , Mohopada".				
4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.				
5) Equipment Used for calibration are calibrated & traceable to national & International standard				
Calibrated By		Approved By		
Calibration Engineer		Technical Manager		
Name : Sanjay More		Name Santosh Sardar		
TF-39, REV:00		End of page		



MD Technologies

Mob.no-8898370272

Your Honest Service partner

MSME Registered No- MH-18-0098825 / GST No-27EBKPB4746M1Z8
Sales and Service -Autoclave, Bio Safety Cabinet, Laminar Air low, Centrifuge,
Bact. Incubator, BOD Incubator, Stability Chamber, Oven, Deep Freezer, etc.

Email-mdtechnologies21@gmail.com

Reg. Office- B54, Shivdham Society, Lane No-4B, New Link Road ,Borivali-west, Mumbai-400103

CUSTOMER:- SRI O Dr. Avinash phadke Mulund DATE:- 06/06/2022

DEPT:- lab

CALL TYPE- AMC/Visit

ADDRESS:- Mulund - (W)

PM Done- 06/06/2022

Mumbai

PM Due.- 05/12/2022

No.	REF	INSTRUMENT MAKE	MODEL	SERIAL NO./ID NO.
1.	Microscope	labomed	Vision 2000	Muld-32 / 140771919
2.				
3.				
4.				

NATURE OF WORK:- Routine Service & PM.

ACTION TAKEN:- Carried out Routine Service


Clean and checked Instrument found ok.
clean its Eye piece, lens Condenser &
optics. Checked with Electrical parts
found working ok

ACTION PENDING:-


SPARES USED FOR REPAIR:-

CUSTOMER S COMMENT:-

ENGINEER S COMMENT:- PM done

 Suresh Bhargava

ENGINEER S NAME & SIGNATURE

 Poojanka Choral
SEAL CUSTOMER S NAME