




ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|---------------------------------|---|-------------------------|---------------|----------------|
| CERTIFICATE NO. | : 220711A/N-101 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 11/07/2022 | Ref. Std. | : ISO:8655:6 | |
| CALIBRATION DUE. | : 10/01/2023 | C.P. No. | : ACS/CP/08 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000067F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Micro Pipette | | | |
| | Make | : Fine Care Corporation | Range | : (2 to 20) µl |
| | L. Count | : | ID No. | : RCL/PIP-01 |
| | Location | : Haematology | S. No. | : QG602677 |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} | | | |

| STANDARD INSTRUMENTS USED : | | | | |
|-----------------------------|-----------------------|--------------|-----------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Weighing Balance | ACS/DB/02/ D449927064 | BlueBox | BBT/223/JAN/22 | 06/01/2023 |

| RESULTS : | | | | |
|-----------------------|-----------------|--------------------|----------------------|------------|
| Mechanical Discipline | | | | |
| Nominal Capacity (µl) | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| 20 | | 10 | 10.08 | (-) 0.08 |
| | | 20 | 20.16 | (-) 0.16 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....


 Auth. Signatory
 S.K. Sharma
 CEO/TM



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | |
|---------------------------------|-------------------------------------|------------------------|--------------|
| CERTIFICATE NO. | : 220711A/N-102 | Page | : 1 of 1 |
| CALIBRATED ON. | : 11 / 07 / 2022 | Ref. Std. | : ISO:8655:6 |
| CALIBRATION DUE. | : 10 / 01 / 2023 | C.P. No. | : ACS/CP/08 |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. |
| ULR No. & ISSUE DATE | : CC342722000000068F Dt. 14.07.2022 | Cal. At | : Site / Lab |

| | |
|-----------------------|---|
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 |
|-----------------------|---|

| | |
|------------------------|--------------------------|
| ITEM CALIBRATED | : Micro Pipette |
| Make | : Fine Care Corporation |
| L. Count | : Range : (10 to 100) µl |
| Location | : ID No. : RCL/PIP-02 |
| | : S. No. : QF602118 |

| | |
|-------------------------------|--|
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} |
|-------------------------------|--|

| STANDARD INSTRUMENTS USED : | | | | |
|-----------------------------|-----------------------|--------------|-----------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Weighing Balance | ACS/DB/02/ D449927064 | BlueBox | BBT/223/JAN/22 | 06/01/2023 |

| RESULTS : | | | | |
|-----------------------|-----------------|--------------------|----------------------|------------|
| Mechanical Discipline | | | | |
| Nominal Capacity (µl) | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| 100 | | 10 | 10.10 | (-) 0.10 |
| | | 50 | 50.20 | (-) 0.20 |
| | | 100 | 100.32 | (-) 0.32 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.


***U.C. :** Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Confirmity statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
CEO/ITM

....End of Calibration Certificate....



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|---------------------------|-----------------------------|--------------------|
| CERTIFICATE NO. | : 220711A/N-103 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 11/07/2022 | Ref. Std. | : ISO:8655:6 | |
| CALIBRATION DUE. | : 10/01/2023 | C.P. No. | : ACS/CP/08 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000069F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Micro Pipette | | | |
| | Make | : Fine Care Corporation | Range | : (100 to 1000) µl |
| | L. Count | : | ID No. | : RCL/PIP-03 |
| | Location | : Haematology | S. No. | : QF601922 |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. { 40% to 60% } | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Weighing Balance | ACS/DB/02/ D449927064 | BlueBox | BBT/223/JAN/22 | 06/01/2023 |
| RESULTS : | | | | |
| Mechanical Discipline | | | | |
| Nominal Capacity (µl) | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| | | 100 | 100.36 | (-) 0.36 |
| 1000 | | 500 | 501.60 | (-) 1.60 |
| | | 1000 | 1003.24 | (-) 3.24 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....






ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|---------------------------------|---|-------------------------|---------------|----------------|
| CERTIFICATE NO. | : 220711A/N-104 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 11 / 07 / 2022 | Ref. Std. | : ISO:8655:6 | |
| CALIBRATION DUE. | : 10 / 01 / 2023 | C.P. No. | : ACS/CP/08 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000070F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Micro Pipette | | | |
| | Make | : Fine Care Corporation | Range | : (2 to 20) µl |
| | L. Count | : | ID No. | : RCL/PIP-04 |
| | Location | : Clinical Biochemistry | S. No. | : QG602682 |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C} R.H. {40% to 60%} | | | |

| STANDARD INSTRUMENTS USED : | | | | |
|-----------------------------|-----------------------|--------------|-----------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Weighing Balance | ACS/DB/02/ D449927064 | BlueBox | BBT/223/JAN/22 | 06/01/2023 |

| RESULTS : | | | | |
|-----------------------|-----------------|--------------------|----------------------|------------|
| Mechanical Discipline | | | | |
| Nominal Capacity (µl) | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| | | 10 | 10.09 | (-) 0.09 |
| 20 | | 20 | 20.14 | (-) 0.14 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmy statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
CEO/TM

....End of Calibration Certificate....



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | |
|---------------------------------|---|------------------------|-------------------------|
| CERTIFICATE NO. | : 220711A/N-105 | Page | : 1 of 1 |
| CALIBRATED ON. | : 11 / 07 / 2022 | Ref. Std. | : ISO:8655:6 |
| CALIBRATION DUE. | : 10 / 01 / 2023 | C.P. No. | : ACS/CP/08 |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. |
| ULR No. & ISSUE DATE | : CC342722000000071F Dt. 14.07.2022 | Cal. At | : Site / Lab |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | |
| ITEM CALIBRATED | : Micro Pipette | Make | : Fine Care Corporation |
| | | Range | : (10 to 100) µl |
| | | L. Count | : ID No. : RCL/PIP-05 |
| | | Location | : S. No. : QF602021 |

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. {40% to 60%}

| STANDARD INSTRUMENTS USED : | | | | |
|-----------------------------|-----------------------|--------------|-----------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Weighing Balance | ACS/DB/02/ D449927064 | BlueBox | BBT/223/JAN/22 | 06/01/2023 |

| RESULTS : | | | | |
|-----------------------|-----------------|--------------------|----------------------|------------|
| Mechanical Discipline | | | | |
| Nominal Capacity (µl) | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| 100 | | 10 | 10.12 | (-) 0.12 |
| | | 50 | 50.26 | (-) 0.26 |
| | | 100 | 100.38 | (-) 0.38 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmation statement not provided since not required.

Prepared by

....End of Calibration Certificate....






ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | |
|---------------------------------|-------------------------------------|------------------------|--------------|
| CERTIFICATE NO. | : 220711A/N-106 | Page | : 1 of 1 |
| CALIBRATED ON. | : 11/07/2022 | Ref. Std. | : ISO:8655:6 |
| CALIBRATION DUE. | : 10/01/2023 | C.P. No. | : ACS/CP/08 |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. |
| ULR No. & ISSUE DATE | : CC342722000000072F Dt. 14.07.2022 | Cal. At | : Site / Lab |

| | |
|-----------------------|---|
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 |
|-----------------------|---|

| | | | |
|------------------------|-------------------------|--------|--------------------|
| ITEM CALIBRATED | : Micro Pipette | | |
| Make | : Fine Care Corporation | Range | : (100 to 1000) µl |
| L. Count | : | ID No. | : RCL/PIP-06 |
| Location | : Clinical Biochemistry | S. No. | : QF601785 |

| | |
|-------------------------------|--|
| ENVIRONMENT CONDITION: | Temp. { 25°C + 3°C } R.H. {40% to 60%} |
|-------------------------------|--|

| STANDARD INSTRUMENTS USED : | | | | |
|-----------------------------|-----------------------|--------------|-----------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Weighing Balance | ACS/DB/02/ D449927064 | BlueBox | BBT/223/JAN/22 | 06/01/2023 |

| RESULTS : Mechanical Discipline | | | | |
|---------------------------------|-----------------|--------------------|----------------------|------------|
| Nominal Capacity (µl) | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| | | 100 | 100.40 | (-) 0.40 |
| 1000 | | 500 | 501.52 | (-) 1.52 |
| | | 1000 | 1003.26 | (-) 3.26 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is $\pm 0.6 \mu\text{l}$ at coverage factor $k=2$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....





ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|-------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-107 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 11 / 07 / 2022 | Ref. Std. | : CG-11 | |
| CALIBRATION DUE. | : 10 / 01 / 2023 | C.P. No. | : ACS/CP/07 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000073F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Digital Thermo Hygrometer | | | |
| | Make : HTC | Range | : (0 to 50)°C, (10-99)% | |
| | L. Count | ID No. | : RCL/LAB/HGM-01 | |
| | Location : Haematology | S. No. | : | |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| RH & Temp. Sensor with Indicator. | ACS/RHTIS/01/ 20365872 | DVG | CC210121000001169F | 15/07/2022 |
| RESULTS | : Thermal Discipline | | | |
| * U. C. INSTT. READS | STANDARD READING | | ERROR | |
| (°C) | 25 | 25.2 | - 0.2 | |
| | 28 | 28.4 | - 0.4 | |
| | 32 | 32.6 | - 0.6 | |
| (%) | 40 | 39 | 1 | |
| | 50 | 49 | 1 | |
| | 60 | 58 | 2 | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4°C, 1.5% at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
CEO/TM




ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | |
|---------------------------------|---|------------------------|-------------------------|
| CERTIFICATE NO. | : 220711A/N-108 | Page | : 1 of 1 |
| CALIBRATED ON. | : 11 / 07 / 2022 | Ref. Std. | : CG-11 |
| CALIBRATION DUE. | : 10 / 01 / 2023 | C.P. No. | : ACS/CP/07 |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. |
| ULR No. & ISSUE DATE | : CC34272200000074F Dt. 14.07.2022 | Cal. At | : Site / Lab |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | |
| ITEM CALIBRATED | : Digital Thermo Hygrometer | | |
| | Make : HTC | Range | : (0 to 50)°C, (10-99)% |
| | L. Count : | ID No. | : RCL/LAB/HGM-02 |
| | Location : Clinical Biochemistry | S. No. | : |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} | | |

| STANDARD INSTRUMENTS USED : | | | | |
|----------------------------------|------------------------|--------------|-------------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| RH & Temp. Sensor with Indicator | ACS/RHTIS/01/ 20365872 | DVG | CC21012100001169F | 15/07/2022 |

| RESULTS : Thermal Discipline | | | |
|------------------------------|------------------|------|-------|
| * U. C. INSTT. READS | STANDARD READING | | ERROR |
| (°C) | 25 | 25.2 | - 0.2 |
| | 28 | 28.2 | - 0.2 |
| | 32 | 32.4 | - 0.4 |
| (%) | 40 | 38 | 2 |
| | 50 | 48 | 2 |
| | 60 | 57 | 3 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4°C, 1.5% at coverage Factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
CEO/ITM

....End of Calibration Certificate....



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-109 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : IS:12508:1988 | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/10 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000075F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Centrifuge | | | |
| | Make : Remi | Range | : (0 to 5250) RPM | |
| | L. Count | ID No. | : RCL/LAB/Cent-01 | |
| | Location : Sample Receiving Room | S. No. | : ZHFN-22095 | |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Tachometer | ACS/DTM/01/ TM5010EK-039 | Godrej | M-210914-16-1 | 14.09.2022- |
| RESULTS | : Mechanical Discipline | | | |
| * U. C. INSTT. READS (RPM) | Standard Measured (RPM) | Error (RPM) | | |
| 1000 | 1003.80 | - 3.80 | | |
| 2000 | 2005.12 | - 5.12 | | |
| 3000 | 3006.48 | - 6.48 | | |
| 4000 | 4006.86 | - 6.86 | | |
| (average of readings) | | | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
GEO/TM


...End of Calibration Certificate...



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-110 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : IS:12508:1988 | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/10 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000076F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Centrifuge | | | |
| | Make : Remi | Range | : (0 to 5250) RPM | |
| | L. Count : | ID No. | : RCL/LAB/Cent-02 | |
| | Location : Sample Receiving Room | S. No. | : ZHFN-18244 | |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Tachometer | ACS/DTM/01/ TM5010EK-039 | Godrej | M-210914-16-1 | 14.09.2022 |
| RESULTS | : Mechanical Discipline | | | |
| * U. C. INSTT. READS (RPM) | Standard Measured (RPM) | Error (RPM) | | |
| 1000 | 1003.78 | - 3.78 | | |
| 2000 | 2005.14 | - 5.14 | | |
| 3000 | 3006.50 | - 6.50 | | |
| 4000 | 4006.88 | - 6.88 | | |
| (average of readings) | | | | |

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....






ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-111 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12/07/2022 | Ref. Std. | : IS:12508:1988 | |
| CALIBRATION DUE. | : 11/01/2023 | C.P. No. | : ACS/CP/10 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000077F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Centrifuge | | | |
| | Make : Remi | Range | : (0 to 6000) RPM | |
| | L. Count : | ID No. | : RCL/LAB/Cent-04 | |
| | Location : Sample Receiving Room | S. No. | : ZIBN-8837 | |
| ENVIRONMENT CONDITION: | Temp. { 25°C + 3°C} R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Tachometer | ACS/DTM/01/ TM5010EK-039 | Godrej | M-210914-16-1 | 14.09.2022 |
| RESULTS | : Mechanical Discipline | | | |
| * U. C. INSTT. READS (RPM) | Standard Measured (RPM) | Error (RPM) | | |
| 1000 | 1003.82 | - 3.82 | | |
| 2000 | 2005.10 | - 5.10 | | |
| 3000 | 3006.52 | - 6.52 | | |
| 4000 | 4006.90 | - 6.90 | | |
| (average of readings) | | | | |

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 3.88 rpm of reading at coverage Factor $k=2$.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....





ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-------------------|
| CERTIFICATE NO. | : 220711A/N-112 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : IS:10996 :1984 | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/09 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC34272200000078F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Digital Timer | | | |
| | Make | : Loudbeeper | Range | : 99 Min |
| | L. Count | : | ID No. | : RCL/LAB/D.TM-01 |
| | Location | : Haematology | S. No. | : |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Timer /Stop Watch Calibrator | ACS/TIM/01/1108J08-038 | Godrej | M-210914-16-2 | 18.09.2022 |
| RESULTS | : ET Discipline | | | |
| * U.C.INSTT. READS (s) | STANDARD READING (s) | ERROR (s) | | |
| 20 | 20.08 | - 0.08 | | |
| 40 | 40.12 | - 0.12 | | |
| 60 | 60.18 | - 0.18 | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.5 s at coverage Factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....





ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-113 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : CG-11 | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/07 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000079F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Digital Temp. Indicator of Refrigerator | | | |
| | Make : Apteckdeals | Range | : (2 to 8) °C | |
| | L. Count : | ID No. | : RCL/LAB/TEMP-01 | |
| | Location : Clinical Biochemistry Fridge | S. No. | : | |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C} R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2021/0784 | 11/11/2022 |
| RESULTS | : ET Discipline | | | |
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) | | |
| 2.06 | 2 | 0.06 | | |
| 4.12 | 4 | 0.12 | | |
| 6.16 | 6 | 0.16 | | |
| 8.22 | 8 | 0.22 | | |
| (average of readings) | | | | |

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....






ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | |
|---------------------------------|-------------------------------------|------------------------|--------------|
| CERTIFICATE NO. | : 220711A/N-114 | Page | : 1 of 1 |
| CALIBRATED ON. | : 12/07/2022 | Ref. Std. | : CG-11 |
| CALIBRATION DUE. | : 11/01/2023 | C.P. No. | : ACS/CP/07 |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. |
| ULR No. & ISSUE DATE | : CC342722000000080F Dt. 14.07.2022 | Cal. At | : Site / Lab |

| | |
|-----------------------|---|
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 |
|-----------------------|---|

| | | | |
|------------------------|---|---------------|-------------------|
| ITEM CALIBRATED | : Digital Temp. Indicator of Refrigerator | | |
| Make | : Apteckdeals | Range | : (2 to 8) °C |
| L. Count | : | ID No. | : RCL/LAB/TEMP-02 |
| Location | : Haematology Fridge | S. No. | : |

| | |
|-------------------------------|--|
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. { 40% to 60% } |
|-------------------------------|--|

| STANDARD INSTRUMENTS USED : | | | | |
|-----------------------------|--------------------|--------------|-----------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2021/0784 | 11/11/2022 |

| RESULTS : ET Discipline | | |
|---------------------------|-----------------------|------------|
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) |
| 2.08 | 2 | 0.08 |
| 4.16 | 4 | 0.16 |
| 6.20 | 6 | 0.20 |
| 8.24 | 8 | 0.24 |
| (average of readings) | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....






ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | |
|---------------------------------|---|------------------------|-------------------|
| CERTIFICATE NO. | : 220711A/N-115 | Page | : 1 of 1 |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : CG-11 |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/07 |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. |
| ULR No. & ISSUE DATE | : CC342722000000081F Dt. 14.07.2022 | Cal. At | : Site / Lab |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | |
| ITEM CALIBRATED | : Digital Temp. Indicator of Refrigerator | | |
| | Make : Apteckdeals | Range | : (2 to 8) °C |
| | L. Count : | ID No. | : RCL/LAB/TEMP-03 |
| | Location : Haematology Freezer | S. No. | : |

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :

| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
|--------------------|--------------------|--------------|-----------------|------------|
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2021/0784 | 11/11/2022 |

RESULTS : ET Discipline

| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) |
|---------------------------|-----------------------|------------|
| 2.10 | 2 | 0.10 |
| 4.14 | 4 | 0.14 |
| 6.18 | 6 | 0.18 |
| 8.26 | 8 | 0.26 |
| (average of readings) | | |

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

...End of Calibration Certificate...

Authn. Signatory
S.K. Sharma
CEO/TM




ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-------------------|
| CERTIFICATE NO. | : 220711A/N-116 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : CG-11 | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/07 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC34272200000082F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Digital Temp. Indicator of Refrigerator | | | |
| | Make | : Apteckdeals | Range | : (2 to 8) °C |
| | L. Count | : | ID No. | : RCL/LAB/TEMP-04 |
| | Location | : Immunoassay Fridge | S. No. | : |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C} R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2021/0784 | 11/11/2022 |
| RESULTS | : ET Discipline | | | |
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) | | |
| 2.06 | 2 | 0.06 | | |
| 4.10 | 4 | 0.10 | | |
| 6.16 | 6 | 0.16 | | |
| 8.20 | 8 | 0.20 | | |
| (average of readings) | | | | |

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....





ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|-------------------------|------------------------|-------------------|
| CERTIFICATE NO. | : 220711A/N-117 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : CG-11 | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/07 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC34272200000083F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Digital Temp. Indicator of Refrigerator | | | |
| | Make | : Apteckdeals | Range | : (2 to 8) °C |
| | L. Count | : | ID No. | : RCL/LAB/TEMP-05 |
| | Location | : Sample Storage Fridge | S. No. | : |
| ENVIRONMENT CONDITION: | Temp. { 25°C + 3°C } R.H. { 40% to 60% } | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2021/0784 | 11/11/2022 - |
| RESULTS | : ET Discipline | | | |
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) | | |
| 2.08 | 2 | 0.08 | | |
| 4.12 | 4 | 0.12 | | |
| 6.20 | 6 | 0.20 | | |
| 8.24 | 8 | 0.24 | | |
| (average of readings) | | | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is $\pm 1.00^\circ\text{C}$ at coverage Factor $k=2$.


***U.C.** : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....






ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-118 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : CG-11 | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/07 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000084F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Digital Temp. Indicator of Refrigerator | | | |
| | Make : Apteckdeals | Range | : (2 to 8) °C | |
| | L. Count : | ID No. | : RCL/LAB/TEMP-06 | |
| | Location : Accessioning Dept. | S. No. | : | |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2021/0784 | 11/11/2022 |
| RESULTS : ET Discipline | | | | |
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) | | |
| 2.10 | 2 | 0.10 | | |
| 4.16 | 4 | 0.16 | | |
| 6.22 | 6 | 0.22 | | |
| 8.28 | 8 | 0.28 | | |
| (average of readings) | | | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....





ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | |
|--|----------------------------------|
| CERTIFICATE NO. : 220711A/N-119 | Page : 1 of 1 |
| CALIBRATED ON. : 12 / 07 / 2022 | Ref. Std. : IS:12508:1988 |
| CALIBRATION DUE. : 11 / 01 / 2023 | C.P. No. : ACS/CP/10 |
| SRF No. & DATE : 220711A/N Dt. 11.07.2022 | Phy. Inspection : O.K. |
| ULR No. & ISSUE DATE : CC34272200000085F Dt. 14.07.2022 | Cal. At : Site / Lab |

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
House no. 40-2, Vivek Vihar, Dehradun- 248001

| | | |
|--|----------------------------|--------------------------------|
| ITEM CALIBRATED : RPM of Shaker | Make : Remi | Range : (80 to 180) RPM |
| | L. Count : | ID No. : RL/LAB/RPM-01 |
| | Location : Serology | S. No. : ZHFS-21343 |

ENVIRONMENT CONDITION: Temp. { 25°C + 3°C } R.H. { 40% to 60% }

| STANDARD INSTRUMENTS USED : | | | | |
|-----------------------------|--------------------------|--------------|-----------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Tachometer | ACS/DTM/01/ TM5010EK-039 | Godrej | M-210914-16-1 | 14.09.2022 |

| RESULTS : Mechanical Discipline | | | |
|---------------------------------|-------------------------|-------------|--|
| * U. C. INSTT. READS (RPM) | Standard Measured (RPM) | Error (RPM) | |
| 80 | 80.80 | - 0.80 | |
| 100 | 101.26 | - 1.26 | |
| 150 | 152.32 | - 2.32 | |
| 180 | 184.68 | - 4.68 | |
| (average of readings) | | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.49 rpm of reading at coverage Factor $k=2$.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Confirmary statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
CEO/TM

....End of Calibration Certificate....



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-120 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : CG-11 | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/07 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000086F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Digital Temp. Indicator of Incubator | | | |
| | Make : Tanco | Range | : (0 to 110) °C | |
| | L. Count : | ID No. | : RL/LAB/INC-01 | |
| | Location : Laboratory | S. No. | : 2K2108013 | |
| ENVIRONMENT CONDITION: | Temp. { 25°C + 3°C } R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2021/0784 | 11/11/2022 |
| RESULTS | : ET Discipline | | | |
| * U.C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) | | |
| 30.70 | 30.0 | 0.70 | | |
| 35.90 | 35.0 | 0.90 | | |
| 41.40 | 40.0 | 1.40 | | |
| 61.60 | 60.0 | 1.60 | | |
| (average of readings) | | | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is $\pm 1.00^\circ\text{C}$ at coverage Factor $k=2$.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

Auth. Signatory
S.K. Sharma
CEO/TM




ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-121 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | | | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | | | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : TDS Meter | | | |
| | Make : HM | Range | : (0 to 1990) ppm | |
| | L. Count : | ID No. | : RCL/LAB/TDS-01 | |
| | Location : Clinical Biochemistry | S. No. | : | |
| ENVIRONMENT CONDITION: | Temp. { 25°C + 3°C } R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital TDS Meter | TDS-01 | ITCCS | ITCCS/2021/21111003 | 09.11.2022 |
| RESULTS | : | | | |
| * U.C. INSTT. READS (ppm) | STANDARD READING (ppm) | ERROR (ppm) | | |
| 302 | 300 | 02 | | |
| 1004 | 1000 | 04 | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is $\pm 1\%$ at coverage Factor $k=2$.


*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Confirmary statement not provided since not required.

Prepared by

....End of Calibration Certificate....






ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-122 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | | | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | | | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : TDS Meter of RO System | | | |
| | Make : Zero B | Range | : (0 to 1990) ppm | |
| | L. Count : | ID No. | : RL/LAB/TDS-01 | |
| | Location : Clinical Biochemistry | S. No. | : NU01183103220002 | |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. { 40% to 60% } | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No./ S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital TDS Meter | TDS-01 | ITCCS | ITCCS/2021/21111003 | 09.11.2022 |
| RESULTS : | | | | |
| * U.C. INSTT. READS (ppm) | STANDARD READING (ppm) | ERROR (ppm) | | |
| 301 | 300 | 01 | | |
| 1003 | 1000 | 03 | | |

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1% at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

