

**Analyte Methodwise Result Report**

Lab : 650  
Lab Name : DR.KAUSHIK PATH LABS  
Report Date : 23/05/2022

Cycle No : 220104  
Sample No : 2  
Sample Date : 26-04-2022

**Biochemistry, Albumin, g/dL**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	288	5.41	0.52	9.59	18
Bromcresol Green (BCG) - (Your Method)	283	5.42	0.52	9.50	18
Agappe Mispa ACE - (Your Analyzer)	3	5.19	0.52	10.05	--

<b>Your Result</b>	5.58	<b>SDI / z-Score</b>	1.50
<b>Assigned Value</b>	4.94	<b>Average SDI</b>	0.82
<b>SDPA</b>	0.428	<b>Uncertainty of AV</b>	0.19
<b>Standard Unit Result Val</b>	5.58	<b>%DEV</b>	2.90

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Albumin**

		N	Mean	SD	CV%	Excluded
	All Analyzer	288	5.41	0.52	9.59	18
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>5.19</b>	<b>0.52</b>	<b>10.05</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	35	5.30	0.25	4.72	2
	Biosystems A15	29	5.85	0.53	9.02	7
	Roche cobas c 311	28	5.71	0.18	3.19	--
	Erba EM 200	27	4.97	0.35	7.00	1
	Erba EM 200	26	4.96	0.35	7.07	1
	Fuji Dri-Chem 4000i	17	5.38	0.28	5.15	--
	ERBA XL-640	16	4.84	0.24	4.86	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650  
Lab Name : DR.KAUSHIK PATH LABS  
Report Date : 23/05/2022

Cycle No : 220104  
Sample No : 2  
Sample Date : 26-04-2022

**Biochemistry, ALP, U/L**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	290	45.39	9.03	19.90	20
IFCC - (Your Method)	33	42.54	10.09	23.71	5
Agappe Mispa ACE - (Your Analyzer)	3	55.37	25.54	46.13	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	49.90	SDI / z' Score	0.54
Assigned Value	42.50	Average SDI	0.34
SDPA	13.700	Uncertainty of AV	4.13
Standard Unit Result Val	49.90	%DEV	17.29


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, ALP		N	Mean	SD	CV%	Excluded
	All Analyzer	290	45.39	9.03	19.90	20
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>55.37</b>	<b>25.54</b>	<b>46.13</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	38	49.35	7.51	15.22	4
	Biosystems A15	29	44.53	8.24	18.50	3
	Roche cobas c 311	28	42.99	2.99	6.95	--
	Erba EM 200	25	43.22	7.01	16.23	2
	Erba EM 200	24	43.18	7.18	16.62	2
	Fuji Dri-Chem 4000i	17	47.44	7.90	16.65	1
	ERBA XL-640	16	41.75	4.11	9.84	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650  
Lab Name : DR.KAUSHIK PATH LABS  
Report Date : 23/05/2022

Cycle No : 220104  
Sample No : 2  
Sample Date : 26-04-2022

**Biochemistry, ALT, U/L**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	292	8.45	2.34	27.72	31
IFCC-Kinetic - (Your Method)	38	9.83	2.60	26.40	7
Agappe Mispa ACE - (Your Analyzer)	3	8.20	1.28	15.62	1


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	7.29	SDI / z' Score	-1.02
Assigned Value	9.43	Average SDI	-0.67
SDPA	2.100	Uncertainty of AV	0.45
Standard Unit Result Val	7.29	%DEV	-25.85


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, ALT		N	Mean	SD	CV%	Excluded
	All Analyzer	292	8.45	2.34	27.72	31
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>8.20</b>	<b>1.28</b>	<b>15.62</b>	<b>1</b>
	Fuji Dri-Chem NX500/i	37	10.19	1.00	9.77	--
	Biosystems A15	29	9.62	2.80	29.14	5
	Roche cobas c 311	28	7.11	1.19	16.71	--
	Erba EM 200	27	8.80	3.02	34.33	4
	Erba EM 200	26	8.96	2.99	33.32	4
	Fuji Dri-Chem 4000i	17	10.12	1.17	11.52	--
	ERBA XL-640	16	8.59	2.48	28.86	3

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650  
Lab Name : DR.KAUSHIK PATH LABS  
Report Date : 23/05/2022

Cycle No : 220104  
Sample No : 2  
Sample Date : 26-04-2022

**Biochemistry, AST, U/L**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	285	15.91	3.73	23.46	13
IFCC-Kinetic - (Your Method)	36	17.65	5.09	28.86	2
Agappe Mispa ACE - (Your Analyzer)	3	19.47	6.43	33.01	--

<b>Your Result</b>	21.80	<b>SDI / z-Score</b>	1.86
<b>Assigned Value</b>	15.50	<b>Average SDI</b>	2.42
<b>SDPA</b>	3.388	<b>Uncertainty of AV</b>	1.28
<b>Standard Unit Result Val</b>	21.80	<b>%DEV</b>	23.51

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, AST		N	Mean	SD	CV%	Excluded
	All Analyzer	285	15.91	3.73	23.46	13
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>19.47</b>	<b>6.43</b>	<b>33.01</b>	--
	Fuji Dri-Chem NX500/i	37	16.58	0.97	5.86	--
	Biosystems A15	28	17.75	5.10	28.74	3
	Erba EM 200	26	17.96	3.84	21.38	3
	Roche cobas c 311	26	12.91	1.96	15.16	--
	Erba EM 200	25	17.95	3.93	21.89	3
	Fuji Dri-Chem 4000i	17	16.29	0.99	6.05	--
	Roche cobas c 111	17	13.16	1.42	10.76	--

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error





**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, Bilirubin Direct, mg/dl**


**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	286	0.52	0.72	137.50	14
Diazotization - (Your Method)	219	0.58	0.81	139.52	11
Agappe Mispa ACE - (Your Analyzer)	3	0.15	0.09	60.13	--

<b>Your Result</b>	0.10	<b>SDI / z' Score</b>	-2.96
<b>Assigned Value</b>	0.46	<b>Average SDI</b>	-2.96
<b>SDPA</b>	0.120	<b>Uncertainty of AV</b>	0.01
<b>Standard Unit Result Val</b>	0.10	<b>%DEV</b>	-82.67

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Bilirubin Direct		N	Mean	SD	CV%	Excluded
	All Analyzer	286	0.52	0.72	137.50	14
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>0.15</b>	<b>0.09</b>	<b>60.13</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	36	0.11	0.02	22.64	2
	Biosystems A15	29	2.23	0.92	41.35	4
	Erba EM 200	27	0.29	0.08	27.84	1
	Roche cobas c 311	27	0.45	0.08	17.30	--
	Erba EM 200	26	0.30	0.08	26.35	1
	Fuji Dri-Chem 4000i	17	0.10	0.00	0.00	--
	ERBA XL-640	16	0.31	0.07	22.93	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, Bilirubin Total, mg/dl**


**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	287	0.63	0.61	97.60	28
TAB - (Your Method)	5	0.57	0.10	16.78	--
Agappe Mispa ACE - (Your Analyzer)	3	0.60	0.08	12.44	--

<b>Your Result</b>	0.68	<b>SDI / z' Score</b>	2.18
<b>Assigned Value</b>	0.44	<b>Average SDI</b>	2.18
<b>SDPA</b>	0.110	<b>Uncertainty of AV</b>	0.03
<b>Standard Unit Result Val</b>	0.68	<b>%DEV</b>	20.14

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Bilirubin Total		N	Mean	SD	CV%	Excluded
	All Analyzer	287	0.63	0.61	97.60	28
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>0.60</b>	<b>0.08</b>	<b>12.44</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	37	0.12	0.05	39.52	3
	Biosystems A15	29	2.13	0.50	23.21	1
	Roche cobas c 311	29	0.50	0.03	6.65	--
	Erba EM 200	26	0.43	0.12	26.91	1
	Erba EM 200	25	0.44	0.12	26.90	1
	Fuji Dri-Chem 4000i	17	0.11	0.04	30.97	2
	ERBA XL-640	15	0.51	0.12	23.23	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650  
Lab Name : DR.KAUSHIK PATH LABS  
Report Date : 23/05/2022

Cycle No : 220104  
Sample No : 2  
Sample Date : 26-04-2022

**Biochemistry, Calcium, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	210	7.23	0.57	7.90	4
Arsenazo III - (Your Method)	75	7.42	0.59	7.95	2
Agappe Mispa ACE - (Your Analyzer)	3	7.80	0.28	3.63	1


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	8.00	SDI / z-Score	0.12
Assigned Value	7.90	Average SDI	-0.07
SDPA	0.834	Uncertainty of AV	0.27
Standard Unit Result Val	8.00	%DEV	7.85


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



PC-1022

**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Calcium**

		N	Mean	SD	CV%	Excluded
	All Analyzer	210	7.23	0.57	7.90	4
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>7.80</b>	<b>0.28</b>	<b>3.63</b>	<b>1</b>
	Fuji Dri-Chem NX500/i	35	6.98	0.42	5.97	--
	Roche cobas c 311	26	7.10	0.27	3.80	1
	Fuji Dri-Chem 4000i	18	7.14	0.38	5.34	--
	Fuji Dri-Chem 4000i	17	7.12	0.38	5.40	--
	Roche cobas c 111	15	7.15	0.32	4.44	--
	Erba EM 200	14	7.56	0.79	10.48	--
	Erba EM 200	13	7.55	0.82	10.91	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



PC-1022

**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, Cholesterol, mg/dl**


**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	285	134.16	10.66	7.94	11
CHOD-PAP - (Your Method)	61	135.14	9.54	7.06	2
Agappe Mispa ACE - (Your Analyzer)	3	132.87	6.87	5.17	--

<b>Your Result</b>	140.80	<b>SDI / z' Score</b>	0.49
<b>Assigned Value</b>	135.73	<b>Average SDI</b>	0.43
<b>SDPA</b>	10.300	<b>Uncertainty of AV</b>	0.69
<b>Standard Unit Result Val</b>	140.80	<b>%DEV</b>	4.19

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Cholesterol		N	Mean	SD	CV%	Excluded
	All Analyzer	285	134.16	10.66	7.94	11
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>132.87</b>	<b>6.87</b>	<b>5.17</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	35	127.85	7.30	5.71	1
	Biosystems A15	28	136.41	11.03	8.08	3
	Roche cobas c 311	28	137.85	6.30	4.57	--
	Erba EM 200	26	132.19	10.42	7.88	2
	Erba EM 200	25	132.33	10.63	8.03	2
	Fuji Dri-Chem 4000i	17	131.78	8.33	6.32	--
	ERBA XL-640	16	129.31	6.57	5.08	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error





**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, Creatinine, mg/dl**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	288	2.89	0.27	9.43	8
Modified Jaffe - (Your Method)	20	2.85	0.23	8.00	3
Agappe Mispa ACE - (Your Analyzer)	3	2.73	0.14	5.21	--

<b>Your Result</b>	2.88	<b>SDI / z' Score</b>	-0.08
<b>Assigned Value</b>	2.90	<b>Average SDI</b>	0.48
<b>SDPA</b>	0.220	<b>Uncertainty of AV</b>	0.02
<b>Standard Unit Result Val</b>	2.88	<b>%DEV</b>	1.02

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Creatinine		N	Mean	SD	CV%	Excluded
	All Analyzer	288	2.89	0.27	9.43	8
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>2.73</b>	<b>0.14</b>	<b>5.21</b>	--
	Fuji Dri-Chem NX500/i	36	2.70	0.13	4.92	--
	Biosystems A15	29	2.97	0.32	10.86	5
	Roche cobas c 311	28	2.96	0.14	4.77	--
	Erba EM 200	27	2.82	0.30	10.49	--
	Erba EM 200	26	2.82	0.30	10.66	--
	Fuji Dri-Chem 4000i	18	2.77	0.13	4.83	--
	Roche cobas c 111	16	2.87	0.16	5.57	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



PC-1022

**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, Glucose, mg/dl**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	288	273.36	20.70	7.57	8
GPO-POD - (Your Method)	25	279.23	19.23	6.89	1
Agappe Mispa ACE - (Your Analyzer)	3	288.85	3.75	1.30	--

<b>Your Result</b>	293.10	<b>SDI / z' Score</b>	0.60
<b>Assigned Value</b>	281.20	<b>Average SDI</b>	1.21
<b>SDPA</b>	19.700	<b>Uncertainty of AV</b>	2.03
<b>Standard Unit Result Val</b>	293.10	<b>%DEV</b>	4.97

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Glucose		N	Mean	SD	CV%	Excluded
	All Analyzer	288	273.36	20.70	7.57	8
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>288.85</b>	<b>3.75</b>	<b>1.30</b>	--
	Fuji Dri-Chem NX500/i	35	253.58	10.47	4.13	--
	Biosystems A15	29	276.96	22.69	8.19	3
	Roche cobas c 311	28	275.14	14.75	5.36	--
	Erba EM 200	26	288.56	17.04	5.90	1
	Erba EM 200	25	288.48	17.40	6.03	1
	Fuji Dri-Chem 4000i	18	252.48	11.47	4.54	--
	ERBA XL-640	16	290.74	14.09	4.85	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, HDL, mg/dl**


**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	266	33.48	6.00	17.91	9
Selective Inhibition Method - (Your Method)	6	33.13	3.19	9.64	--
Agappe Mispa ACE - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	37.80	<b>SDI / z' Score</b>	0.78
<b>Assigned Value</b>	33.68	<b>Average SDI</b>	0.75
<b>SDPA</b>	5.300	<b>Uncertainty of AV</b>	0.63
<b>Standard Unit Result Val</b>	37.80	<b>%DEV</b>	14.11

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, HDL		N	Mean	SD	CV%	Excluded
	All Analyzer	266	33.48	6.00	17.91	9
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	33	32.22	1.79	5.56	1
	Roche cobas c 311	28	29.98	1.44	4.79	--
	Erba EM 200	26	33.99	4.89	14.39	1
	Biosystems A15	25	42.76	11.37	26.60	3
	Erba EM 200	25	34.06	4.98	14.63	1
	ERBA XL-640	16	33.64	3.06	9.08	--
	Fuji Dri-Chem 4000i	15	32.07	1.71	5.32	--
	Roche cobas c 111	15	30.13	1.60	5.32	--

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650  
Lab Name : DR.KAUSHIK PATH LABS  
Report Date : 23/05/2022

Cycle No : 220104  
Sample No : 2  
Sample Date : 26-04-2022

**Biochemistry, LDL, mg/dl**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	98	74.00	6.96	9.41	4
Calculated - (Your Method)	35	75.59	5.78	7.65	3
Agappe Mispa ACE - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	77.50	<b>SDI / z-Score</b>	0.65
<b>Assigned Value</b>	73.83	<b>Average SDI</b>	0.55
<b>SDPA</b>	5.640	<b>Uncertainty of AV</b>	0.94
<b>Standard Unit Result Val</b>	77.50	<b>%DEV</b>	2.53

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, LDL**

	N	Mean	SD	CV%	Excluded
All Analyzer	98	74.00	6.96	9.41	4
<b>Your Analyzer : Agappe Mispa ACE</b>	1	--	--	--	--
Roche cobas c 311	21	73.41	4.07	5.54	1
Erba EM 200	10	77.27	8.71	11.28	1
ERBA XL-640	8	73.25	4.32	5.90	1
Roche cobas 6000	8	70.69	1.87	2.65	--
Roche cobas INTEGRA	8	74.49	6.54	8.78	--
Horiba ABX Pentra C200	5	59.83	3.74	6.25	--
Roche cobas 8000	5	72.13	6.04	8.37	--
Roche cobas c 111	5	73.72	7.09	9.61	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, Potassium, mmol/L**

**Comparative Statistics**



	N	Mean	SD	CV%	Excluded
All Methods	218	3.68	0.17	4.53	4
ISE direct - (Your Method)	159	3.69	0.17	4.53	4
Acculite II - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Method wise Peer Grouping**

Your Result	3.60	SDI / z' Score	-0.33
Assigned Value	3.70	Average SDI	0.16
SDPA	0.310	Uncertainty of AV	0.01
Standard Unit Result Val	3.60	%DEV	-2.36

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Potassium**

	N	Mean	SD	CV%	Excluded
All Analyzer	218	3.68	0.17	4.53	4
<b>Your Analyzer :</b> Acculite II	1	--	--	--	--
Fuji Dri-Chem NX500/i	41	3.72	0.14	3.73	--
Fuji Dri-Chem 4000i	19	3.72	0.17	4.60	--
Roche cobas c 311	19	3.69	0.16	4.26	--
Fuji Dri-Chem 4000i	18	3.70	0.16	4.35	--
Roche AVL 9180	18	3.76	0.12	3.22	--
EasyLyte	10	3.57	0.18	5.05	--
BeneSphera Electrolyte Analyzer	10	3.57	0.15	4.15	1
Bene Sphera C73	8	3.64	0.16	4.40	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, Total Protein, g/dL**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	282	8.23	0.49	5.99	13
Modified Biuret end point - (Your Method)	9	8.39	0.39	4.66	--
Agappe Mispa ACE - (Your Analyzer)	3	8.24	0.62	7.49	--

<b>Your Result</b>	8.83	<b>SDI / z-Score</b>	0.23
<b>Assigned Value</b>	8.69	<b>Average SDI</b>	0.32
<b>SDPA</b>	0.583	<b>Uncertainty of AV</b>	0.19
<b>Standard Unit Result Val</b>	8.83	<b>%DEV</b>	5.19

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Total Protein		N	Mean	SD	CV%	Excluded
	All Analyzer	282	8.23	0.49	5.99	13
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>8.24</b>	<b>0.62</b>	<b>7.49</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	36	8.31	0.29	3.49	2
	Biosystems A15	29	8.33	0.67	8.05	4
	Roche cobas c 311	28	8.12	0.25	3.03	--
	Erba EM 200	25	8.46	0.54	6.37	3
	Erba EM 200	24	8.42	0.52	6.12	3
	Fuji Dri-Chem 4000i	17	8.29	0.31	3.69	--
	ERBA XL-640	16	8.55	0.49	5.67	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, Triglycerides, mg/dl**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	281	114.44	13.60	11.88	7
GPO-TOPS - (Your Method)	6	113.07	15.39	13.61	--
Agappe Mispa ACE - (Your Analyzer)	3	116.80	13.82	11.83	--

<b>Your Result</b>	131.70	<b>SDI / z' Score</b>	1.30
<b>Assigned Value</b>	118.92	<b>Average SDI</b>	1.17
<b>SDPA</b>	9.800	<b>Uncertainty of AV</b>	0.94
<b>Standard Unit Result Val</b>	131.70	<b>%DEV</b>	16.48

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Triglycerides		N	Mean	SD	CV%	Excluded
	All Analyzer	281	114.44	13.60	11.88	7
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>116.80</b>	<b>13.82</b>	<b>11.83</b>	--
	Fuji Dri-Chem NX500/i	34	122.85	7.75	6.31	--
	Biosystems A15	29	104.42	15.30	14.65	2
	Roche cobas c 311	28	119.82	6.75	5.63	--
	Erba EM 200	25	100.00	10.80	10.80	--
	Erba EM 200	24	99.73	10.95	10.98	--
	ERBA XL-640	16	105.23	9.99	9.50	1
	Fuji Dri-Chem 4000i	16	120.75	7.24	6.00	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



**Analyte Methodwise Result Report**

Lab : 650  
Lab Name : DR.KAUSHIK PATH LABS  
Report Date : 23/05/2022

Cycle No : 220104  
Sample No : 2  
Sample Date : 26-04-2022

**Biochemistry, Urea, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	288	18.63	2.03	10.91	20
Urease, UV - (Your Method)	52	18.27	2.51	13.76	2
Agappe Mispa ACE - (Your Analyzer)	3	18.97	1.03	5.41	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	20.10	SDI / z' Score	0.94
Assigned Value	18.62	Average SDI	1.60
SDPA	1.570	Uncertainty of AV	0.16
Standard Unit Result Val	20.10	%DEV	10.05


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Urea		N	Mean	SD	CV%	Excluded
	All Analyzer	288	18.63	2.03	10.91	20
<b>Your Analyzer :</b>	<b>Agappe Mispa ACE</b>	<b>3</b>	<b>18.97</b>	<b>1.03</b>	<b>5.41</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	36	18.96	1.07	5.64	4
	Biosystems A15	29	17.40	1.84	10.57	2
	Roche cobas c 311	28	18.91	1.17	6.16	1
	Erba EM 200	26	20.61	2.15	10.42	1
	Erba EM 200	25	20.79	1.98	9.54	1
	Fuji Dri-Chem 4000i	17	18.84	0.87	4.63	2
	ERBA XL-640	16	18.08	1.06	5.89	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error





**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

**Biochemistry, Uric Acid, mg/dl**

**Comparative Statistics**


	N	Mean	SD	CV%	Excluded
All Methods	286	3.89	0.47	11.97	15
Uricase - POD - (Your Method)	38	4.12	0.74	17.97	4
Agappe Mispa ACE - (Your Analyzer)	3	4.44	0.33	7.50	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	4.82	SDI / z' Score	2.86
Assigned Value	3.62	Average SDI	2.62
SDPA	0.420	Uncertainty of AV	0.03
Standard Unit Result Val	4.82	%DEV	17.02

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 650

Cycle No : 220104

Lab Name : DR.KAUSHIK PATH LABS

Sample No : 2

Report Date : 23/05/2022

Sample Date : 26-04-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Uric Acid**

	N	Mean	SD	CV%	Excluded
All Analyzer	286	3.89	0.47	11.97	15
<b>Your Analyzer : Agappe Mispa ACE</b>	<b>3</b>	<b>4.44</b>	<b>0.33</b>	<b>7.50</b>	<b>--</b>
Fuji Dri-Chem NX500/i	37	4.17	0.21	5.01	--
Biosystems A15	29	4.05	0.39	9.53	3
Roche cobas c 311	26	3.52	0.21	6.04	--
Erba EM 200	25	4.02	0.24	5.85	2
Erba EM 200	24	4.01	0.24	5.98	2
Fuji Dri-Chem 4000i	20	4.08	0.32	7.94	1
ERBA XL-640	16	4.12	0.33	7.91	1

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070



PC-1022