



**TULIP DIAGNOSTICS (P) LTD.**

a PerkinElmer company

CIN: U85195GA1988PTC005717

**Thrissur Office:**

Thattil Nadakalan Complex, Kuriachira P.O.,

Thrissur - 680 006, INDIA.

Tel.: (0487) 2250793, 6544048. Mobile No.: 9645084048.

E-mail: tulipkerala@gmail.com

**08.07.2022**

**CALIBRATION CERTIFICATE**

This is to certify that, CORALAB TOUCH Semi-Automatic Analyzer bearing Serial No **BS8KO195E**

Installed at, M/s **RAJALAKSHMI CLINICAL LABORATORY, PONDY ROAD, MARAKKANAM** is performing satisfactorily.

CORALAB TOUCH is calibrated last on **08.07.2022** by our Service Engineer,


**Mr. JEBARAJAN**

Following checks have been performed:

1. Peristaltic Pump Cleaned
2. Cleaned flowcell & Tubings
3. Adjusted Lamp Gain.
4. Checked all the Voltage.

Next Calibration Due on **07.07.2023**

For **TULIP DIAGNOSTICS (P) LTD**



**AUTHORISED SIGNATORY**

**Corporate Office:** Gitanjali, Tulip Block, Dr. Antonio Do Rego Bagh, Alto Santacruz, Bambolim Complex P.O., Goa - 403 202, INDIA.  
Tel.: (0832) 2458546-51, Fax: (0832) 2458544. E-mail: sales@tulipgroup.com, Website: www.tulipgroup.com

**Innovative Excellence**