



**CERTIFICATE of CALIBRATION**

**Manufacturer : Ark Diagnostics Bangalore Pvt Ltd**

**Instrument : HPLC Analyser - CKK -A1C.**

**Serial Number : 123-027**

**Customer : Mediquest Health Care.**

Ark Diagnostics Bangalore Pvt. Ltd., certifies that the above listed instrument was calibrated on date noted at. Mediquest healthcare, Mahim complex, Pumpwell, old kankanady road, near fether muller circle, kankanady, Mangalore-575002, in accordance with the specifications set forth by the manufacturer using the certified control material as per the provided assay values and expected ranges from the manufacturer.

**Calibration Date:-19<sup>th</sup> January 2022**

**Calibrated by: - Mr. Vineeth.**

**Calibration Due:-18<sup>th</sup> January 2023.**

**For ark Diagnostics Bangalore Pvt. Ltd.,**

**Authorized signature**



Patient report

\*\*\* CKKA1c 1.3.8.S \*\*\*


Date: 2022/07/22 09:41:00

NO: 02484 ID:QC2

Method: HbA1c

HbA1c 09.8 % NGSP ↑  
( 4.0% - 6.0% )

eAG 083.6 mmol/mol IFCC  
234.6 mg/dL



NAME	TIME	AREA	%
A1a	027	0002.5	00.2
A1b	035	0019.0	01.5
F	054	0021.8	01.8
LA1c	075	0004.6	00.4
A1c	088	0080.2	09.8
A0	156	1115.0	89.6

Total Area: 1243.3

*Handwritten signature*  
22/7/22

HbA1c ANALYZER

\*\*\* CKKA1c 1.3.8.S \*\*\*