



REF. NO.AB/TSC/CC/001

Caibration Certificate

| | |
|---------------------------|---|
| Customer Name | ADVANCE MALAVIYA PATHALOGY LABORATORY |
| Model | Fully Automated 3-Parts Celcounter SYSMEX KX-21 |
| Serial no. | B8349 |
| Calibration Date | 04/07/2022 |
| Next Calibration Due Date | 03/07/2023 |

This is to certify that the above mentioned product has been calibrated by calibrator CBC NEK CAL R & D SYSTEM USA LOT NO. NK0722 EXP.05/08/2022 for 5 parameters (WBC,RBC,HGB,MCV and PLT) according to standard procedure provided by Manufacturer. Also varify by Normal control CBC ST PLUSLOT No ST0622N EXP. 05/09/2022 and results are in range provided by R & D system USA.

The refernce instruments used for value assignment are managed by R & D system and these are traceable to the International Standards, such as ICSH.

Calibration performed by,



Service Engineer,

Ap helion Biotech

Surat.

Attachment

1. Certificate of Inspection(TSR).
2. Assay Sheet of Calibrator
3. Printouts

APHELION BIOTECH

BLOCK NO-317, PLOT NO-113, SAHYOG INDUSTRIAL ESTATE, DASTAN-KARELI ROAD,MOJE-KARELI, TA- PALSANA, DIST.-SURAT - 394310,
M. 90990 18292 / 9898202300 E-mail: ap helionbiotech@gmail.com