

VIEW LAB MONTHLY SUMMARY

Lab Name **AVM DIAGNOSTIC CENTRE**

[Details About Robust Analysis](#)

Lab No **3393**

[Detail About Monthly Summary](#)

Month **June**

[Detail about SDI](#)

Year **2022**

Constituent Group **Chemistry II**

[Click on the analyte to view Graphical Data](#)

All Analyser Result

Print

Date of Result Entered : **18/06/2022**

Date of Report Published : **07/07/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Biosystems	280	85.34	9.80	8.36	83 mg/dL	-0.28	1.00
2	UREA II	Urease UV / GLDH II	Biosystems	225	43.22	10.99	4.75	42 mg/dL	-0.26	0.63
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	240	2.00	10.87	0.22	2.1 mg/dL	0.46	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	236	4.09	12.19	0.50	3.87 mg/dL	-0.44	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	258	6.01	7.82	0.47	6.4 g/dL	0.83	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	254	3.61	9.65	0.35	3.7 g/dL	0.26	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	243	7.10	11.67	0.83	7.4 mg/dL	0.36	0.11
8	CHOLESTROL II	CHOD-PAP II	Biosystems	270	124.12	9.71	12.05	128 mg/dL	0.32	1.47
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	249	170.46	9.67	16.49	160 mg/dL	-0.63	2.09

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out