



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 1717

Distribution No.: 156-C

Month/Year: May/2022

Instrument ID: XP-100 (S.NO. B6502)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 12-07-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	21.2	21.1	42.3	41.35	0.2760	0.04	0.1	0.21	0.0140	-0.37
RBC x10 ⁶ /µl	1	4.32	4.3	8.62	8.63	0.0080	-0.04	0.02	0.03	0.0020	-0.27
Hb g/dl	1	13.3	13.3	26.6	26.4	0.0330	0.25	0	0.1	0.0080	-0.67
HCT%	1	40.1	39.8	79.9	83.6	0.2070	-0.60	0.3	0.4	0.0240	-0.27
MCV-fl	1	92.8	92.6	185.4	194.1	0.4040	-0.73	0.2	0.3	0.0250	-0.17
MCH-Pg	1	30.9	30.8	61.7	61.1	0.0840	0.27	0.1	0.2	0.0160	-0.45
MCHC-g/dl	1	33.4	33.2	66.6	62.6	0.1590	0.87	0.2	0.3	0.0170	-0.34
Plt. x10 ³ /µl	1	225	223	448	346	3.20	1.01	2	5	0.31	-0.54
Retic %	2	2.2	1.5	3.7	7.85	0.12	-1.22	0.7	0.3	0.02	1.80

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=0 , Poly=12 L=24, E=1, Mono/Promono=3 , B1=60 P.M.=0, Mye=0, Meta=0, Other=Nil	Blast: 49-70, Lympho: 12-27 ,Poly: 9-17, /mono:1-5 nRBC/Eosino/Myelo/Meta/promyelo: 0-2		
RBC Morphology	3	Normocytic normochromic, few microcytes, anisocytosis (+)	Predominantly: Normocytic/ Normochromic, Moderate: Anisocytosis, Microcytic , macrocytes, Tear drop cells		
Diagnosis	3	Acute leukemia (subleukemic)	Acute Leukemia (AL)		

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 156--C	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	348	346	99.13	89.88	0.29	3.47	0.58	6.65
RBC x10⁶/µl	1	348	348	91.09	90.8	4.6	3.74	4.31	5.46
Hb g/dl	1	348	348	85.92	87.93	6.61	4.6	7.47	7.47
HCT%	1	348	347	88.47	89.05	4.9	3.75	6.63	7.2
MCV-fl	1	348	347	88.47	91.93	3.75	4.61	7.78	3.46
MCH-Pg	1	348	347	87.32	87.61	5.76	6.92	6.92	5.47
MCHC-g/dl	1	348	347	91.64	86.46	4.9	4.9	3.46	8.64
Plt. x10³/µl	1	348	347	95.39	91.64	3.46	3.46	1.15	4.9
ReticCount%	2	348	330	90.91	82.12	7.27	12.73	1.82	5.15
PS Assessment	3	348	327	Satisfactory :94.27%, Borderline Sat. :4.59%, Unsatisfactory :1.14%					

***Comments:**

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----