

**CERTIFICATE OF CALIBRATION**

ADL/DO/MS/C/0145/21-22

Instrument: Mispa CXL PRO Fully Automated Biochemistry Analyser**Customer Name:** Dr. Vaibhav's Pathocity Path Lab & Imaging, Paschim Vihar, New Delhi**Date of Installation:** 27/Aug/2021**S/N:** 210T180CS0218K**Alignment Check**

Motion Check	Checked	Remark
Sample Probe	<input checked="" type="checkbox"/>	OK
Reagent Probe	<input checked="" type="checkbox"/>	OK
Mixing Bar	<input checked="" type="checkbox"/>	OK
Reaction Disk	<input checked="" type="checkbox"/>	OK
Reagent Disk	<input checked="" type="checkbox"/>	OK
Sample Disk	<input checked="" type="checkbox"/>	OK
Cuvette Check	<input checked="" type="checkbox"/>	OK
Lamp	<input checked="" type="checkbox"/>	OK

System	Checked	Remark
Main unit	<input checked="" type="checkbox"/>	OK
Temp. unit	<input checked="" type="checkbox"/>	OK
Mixing unit	<input checked="" type="checkbox"/>	OK
Reaction unit	<input checked="" type="checkbox"/>	OK
Reagent unit	<input checked="" type="checkbox"/>	OK
Sample unit	<input checked="" type="checkbox"/>	OK

AGAPPE DIAGNOSTICS LTD.

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkattoffice@agappe.in

BANGALORE OFFICE
S-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**OPTICAL SECTION**

Filter	Light Intensity	
340	15078	
380	14250	
405	13713	
450	12732	
480	12500	
505	12231	
546	12051	
570	11823	
600	11523	
660	11421	
700	11341	
800	11040	

Power Supply :

12V 24V

The equipment is calibrated for the different parameters by using Multi Calibrator

Lot.No: 31050609 Exp: 2023-09

This is to certify that the above mentioned product has been calibrated according to the standard operating procedure given by the Agappe Diagnostics Ltd.

Date of Calibration : 27/Aug/2021

Next Due : 26/Feb/2022

For Agappe Diagnostics Ltd.

Authorized Signatory

Name: Mukesh Sharma

Designation: Manager - Application Support

**AGAPPE DIAGNOSTICS LTD.**

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaishigh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in