

TO,
THE ADMIN
NABL Medical Entry Level
Testing M(EL)T Labs Program

SUB :- Revert of Queries regarding Clinical Pathology

Respected,
Sir/ Mam

Urine Leucocyte test parameter not included in Metropolis EQAS. It can be Excluded
From Scope.

Thanks



Dr. Kaushik Path Labs



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022

Clinical Path - Urine Routine Chemistry,Bilirubin (urine), -

	N
All Methods	39
Photoelectric colorimetry - (Your Method)	4
Urometer 120 - (Your Analyzer)	3


Evaluation Based On - Method wise Peer Grouping

Your Result Negative

Result Expected Neg

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022



Reported Results Legends :-

+1	: 1+
+2	: 2+
+3	: 3+
Neg	: Negative

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error

Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022

Clinical Path - Urine Routine Chemistry,Blood (urine), -

	N
■ All Methods	40
■ Photoelectric colorimetry - (Your Method)	4
■ Urometer 120 - (Your Analyzer)	3


Evaluation Based On - Method wise Peer Grouping

Your Result 2+

Result Expected +1,+2,Tr

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022



Reported Results Legends :-

+1	: 1+
+2	: 2+
+3	: 3+
+4	: 4+
Tr	: Trace
Neg	: Negative

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error

Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Clinical Path - Urine Routine Chemistry, Glucose (urine), -

	N
■ All Methods	43
■ Photoelectric colorimetry - (Your Method)	4
■ Urometer 120 - (Your Analyzer)	3

Evaluation Based On - Method wise Peer Grouping	
Your Result	4+
Result Expected	+2,+3,+4

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-

+1	: 1+
+2	: 2+
+3	: 3+
+4	: 4+
Tr	: Trace
Neg	: Negative

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022

Clinical Path - Urine Routine Chemistry, Ketone (urine), -


	N
■ All Methods	39
■ Photoelectric colorimetry - (Your Method)	4
■ Urometer 120 - (Your Analyzer)	3

Evaluation Based On - Method wise Peer Grouping

Your Result	Negative
Result Expected	Neg

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022



Reported Results Legends :-

+1	: 1+
+2	: 2+
+3	: 3+
Tr	: Trace
Neg	: Negative

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error

Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Clinical Path - Urine Routine Chemistry,Nitrite (urine), -

	N
■ All Methods	39
■ Photoelectric colorimetry - (Your Method)	4
■ Urometer 120 - (Your Analyzer)	3

Evaluation Based On - Method wise Peer Grouping	
Your Result	Absent
Result Expected	Abs

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022



Reported Results Legends :-
 Abs : Absent
 Pre : Present

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error

Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Clinical Path - Urine Routine Chemistry,pH, -

	N
■ All Methods	40
■ Photoelectric colorimetry - (Your Method)	3
■ Urometer 120 - (Your Analyzer)	3

Evaluation Based On - Method wise Peer Grouping	
Your Result	6
Result Expected	5,5.5,6

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022



Reported Results Legends :-

5	: 5
5.5	: 5.5
6	: 6
6.5	: 6.5
7	: 7
7.5	: 7.5
8	: 8
8.5	: 8.5
9	: 9

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error

Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022

Clinical Path - Urine Routine Chemistry,Protein (Urine), -


	N
■ All Methods	42
■ Photoelectric colorimetry - (Your Method)	5
■ Urometer 120 - (Your Analyzer)	3

Evaluation Based On - Method wise Peer Grouping

Your Result	1+
Result Expected	+1,+2,Tr

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-

+1	: 1+
+2	: 2+
+3	: 3+
Tr	: Trace
Neg	: Negative

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error

Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Clinical Path - Urine Routine Chemistry, Specific Gravity, -

	All Methods	N 40
	Photoelectric colorimetry - (Your Method)	3
	Urometer 120 - (Your Analyzer)	3

Evaluation Based On - Method wise Peer Grouping	
Your Result	1.030
Result Expected	1.020,1.025,1.030

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022



Reported Results Legends :-

1.005	: 1.005
1.010	: 1.010
1.015	: 1.015
1.020	: 1.020
1.025	: 1.025
1.030	: 1.030
1.035	: 1.035

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error



Metropolis Healthcare Ltd., EQAS

Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022


Clinical Path - Urine Routine Chemistry,Urine Pregnancy Test, -

	N
All Methods	30
Immunochromatography - (Your Method)	29
- - (Your Analyzer)	30

Evaluation Based On - Method wise Peer Grouping	
Your Result	Negative
Result Expected	Neg

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022



Reported Results Legends :-
 Pos : Positive
 Neg : Negative

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error

Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Clinical Path - Urine Routine Chemistry,Urobilinogen (urine), -

	N
 All Methods	37
 Photoelectric colorimetry - (Your Method)	4
 Urometer 120 - (Your Analyzer)	3

Evaluation Based On - Method wise Peer Grouping	
Your Result	Normal
Result Expected	Nor

Note : # is an indication that this parameter has not been evaluated



Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-

Nor	: Normal
+1	: 1+
+2	: 2+
+3	: 3+
+4	: 4+

Comments :

- Interpretation when an analyte is reported for multiple samples in a Round:
- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
 - b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error

AK Nigam

