

**SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE  
GURUGRAM**

Sir,

This is in concern of Clinical Pathology EQAS, We are enrolled for Clinical Pathology EQAS with Metropolis Healthcare Ltd and Successfully completed the same last year. We have submitted the EQAS reports of Clinical Pathology dated Dec 2021 during our NABL MELT application submission in May 2022.

We have received only the Clinical Pathology Urine slide from the EQAS provider in the last round and report of the same is submitted to you in Jun 2022.

We are scheduled to get the EQAS Urine samples in next round only.

Kindly accept our submission.

*Visharma*  
*26-07-22*  
Dr Vijay Kumar Sharma  
Senior Manager (Quality)  
HLL Lifecare Limited

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit
# Other findings	-	Squamous Epithelial Cells	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: Parameter other findings not evaluated

Corrective Action: Sufficient peers are not available with EQAS provider for evaluation.

Reviewed by: Visharma

Dated: 23.05.22

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Crystal	-	Triple phosphate crystal	Triple phosphate crystal
@ Casts	-	No cast seen	No cast seen
# Other findings	-	Squamous Epithelial Cells	-

Legend @ : Acceptable  
X : Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	1
Evaluated Parameters	2
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Sediment Analysis	100.00 %

Image

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kirof Road, Kurla (W),  
Mumbai - 400070

*Visharma*  
*23.05.22*

End of Report

# *Certificate of Participation*

**Metropolis Healthcare Ltd. EQAS**

**Cycle 2021 – January to December**

**SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (HLL Lifecare Ltd. Gurgaon)**

Lab ID: 549

Has successfully participated in Scheme for

●M(EL)T ●IMMUNOASSAY ●Clinical Pathology



*Dr Puneet Kumar Nigam*

*Dr Puneet Kumar Nigam*

*PT Coordinator*

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210512 and Sample No 01

Report Date : 30/12/2021

Analyte	Instrument	Result Value	Standard Unit
# Blood (urine)	Uriplus 600	1+	-
# Nitrite (urine)	Uriplus 600	Present	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: Blood and Nitrite was not evaluated by EQAS provider due to less peer submission.

Corrective Action:

Reviewed by:

*Visharma*

Dated:

30-12-21

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Result Details For Cycle No 210512 and Sample No 01

Report Date : 30/12/2021

Instrument : Uriplus 600

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5.5,6,6.5
@ Specific Gravity	-	1.025	1.015,1.020,1.025
@ Glucose (urine)	-	2+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,+2,Tr
# Blood (urine)	-	1+	
# Nitrite (urine)	-	Present	
@ Ketone (urine)	-	Negative	Negative
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal

Legend @ : Acceptable  
X : Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	2
Evaluated Parameters	7
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kiroli Road, Kurla (W),  
Mumbai - 400070

*Visharma*

End of Report

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Corrective Action: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Reviewed by: Visharma

Dated: 19-07-21

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Instrument : Uriplus 600

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	6.6.5
@ Specific Gravity	-	1.020	1.010,1.015,1.020
@ Glucose (urine)	-	2+	+1,+2,+3
@ Protein (Urine)	-	1+	+1,+2
@ Ketone (urine)	-	Negative	Negative
@ Blood (urine)	-	Negative	Negative
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
@ Nitrite (urine)	-	Present	Present

Legend @ : Acceptable  
X : Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kiroli Road, Kurla (W),  
Mumbai - 400070

*Visharma*

End of Report