



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2022



PC-1024

Lab Name **RAASI DIAGNOSTIC CENTRE** Lab No **11073**
 Constituent Group **Chemistry II** Date of Result Entered : **20/06/2022**
 PT item **Lyophilized Serum** Date of Report Published : **07/07/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Biosystems	280	85.34	9.80	8.36	73.3 mg/dL	-1.44	1.00
2	UREA II	Urease UV / GLDH II	Biosystems	225	43.22	10.99	4.75	40.6 mg/dL	-0.55	0.63
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	240	2.00	10.87	0.22	2.13 mg/dL	0.60	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	236	4.09	12.19	0.50	3.85 mg/dL	-0.48	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	258	6.01	7.82	0.47	5.67 g/dL	-0.72	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	254	3.61	9.65	0.35	3.1 g/dL	-1.47	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	243	7.10	11.67	0.83	7.87 mg/dL	0.93	0.11
8	CHOLESTROL II	CHOD-PAP II	Biosystems	270	124.12	9.71	12.05	111.7 mg/dL	-1.03	1.47
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	249	170.46	9.67	16.49	166.9 mg/dL	-0.22	2.09

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

RAASI DIAGNOSTIC CENTRE
130- TRICHY ROAD, NEAR FIRE STATION, THURAIYUR
TRICHY
TAMILNADU621010

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****