

HIN/MED/2021-2022/10271

15 April 2022

CALIBRATION CERTIFICATE

This is to certify that Heamatology analyser **YH500** bearing serial number-**003YOXH03371** installed at **Arogyam Pathology Center, Pai Bhavan, Tonca Caranzalem, Goa** was calibrated on **8th April 2022**.

Calibrator: ABX MINOCAL

Lot: CX465

Expiry: 6/5/2022.

Reports of blank cycle, repeatability and calibration values are found in acceptable range.

Next calibration cycle 7th April 2022

A circular stamp with the text "H. RIBA IN PVT. LTD." around the perimeter. Overlaid on the stamp is a handwritten signature in black ink.

Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**