



Thrissur Office:

Thattil Nadakalan Complex, Kuriachira P.O.,

Thrissur - 680 006, INDIA.

Tel.: (0487) 2250793, 6544048. Mobile No.: 9645084048.

E-mail: tulipkerala@gmail.com

08.07.2022

CALIBRATION CERTIFICATE

This is to certify that, CORALAB TOUCH Semi-Automatic Analyzer bearing Serial No **BS8KO195E**

Installed at, M/s **RAJALAKSHMI CLINICAL LAB, NO 51/2, PONDY ROAD, MARAKKANAM PINCODE 604303** is performing satisfactorily.

CORALAB TOUCH is calibrated last on **08.07.2022** by our Service Engineer,

Mr. JEBARAJAN

Following checks have been performed:

1. Peristaltic Pump Cleaned
2. Cleaned flowcell & Tubings
3. Adjusted Lamp Gain.
4. Checked all the Voltage.

Next Calibration Due on **07.07.2023**

For **TULIP DIAGNOSTICS (P) LTD**



AUTHORISED SIGNATORY

RAJALAKSHMI LAB
Coralab Touch
27-07-2022 11:30:18
Nr: 2018
Ref : glu qc
Program Name: GLU
OD.:0.4817
CONC: 112.030 mg/dL H
Normal Low: 70.0000
Normal High: 110.0000

RAJALAKSHMI LAB
Coralab Touch
27-07-2022 12:11:21
Nr: 2026
Ref : UREA QC
Program Name: UREA
OD.:0.2225
CONC: 27.859 mg/dL
Normal Low: 10.0000
Normal High: 40.0000

RAJALAKSHMI LAB
Coralab Touch
27-07-2022 11:55:07
Nr: 2021
Ref : CREAT QC
Program Name: CREAT
OD1:0.2909
OD2:0.3215
CONC: 1.488 mg/dL H
Normal Low: 0.6000
Normal High: 1.4000

RAJALAKSHMI LAB
Coralab Touch
27-07-2022 13:37:57
Nr: 2041
Ref : TBILL QC
Program Name: TBIL
OD1:-0.0011
OD2:0.1156
CONC: 1.517 mg/dL H
Normal Low: 0.6000
Normal High: 1.4000

RAJALAKSHMI LAB
Coralab Touch
27-07-2022 12:29:14
Nr: 2029
Ref : TPROTIN QC
Program Name: TOTAL PROTEIN
OD.:0.4108
CONC: 5.419 g/dL L
Normal Low: 6.0000
Normal High: 8.0000

RAJALAKSHMI LAB
Coralab Touch
27-07-2022 12:09:17
Nr: 2025
Ref : ALBUMIN QC
Program Name: ALBUMIN
OD.:0.5920
CONC: 2.859 g/dL L
Normal Low: 3.7000
Normal High: 5.3000

RAJALAKSHMI LAB
Coralab Touch
27-07-2022 12:46:07
Nr: 2033
Ref :
Program Name: URIC ACID
OD.:0.0873
CONC: 4.584 mg/dL
Normal Low: 3.2000
Normal High: 7.4000

RAJALAKSHMI LAB
Coralab Touch
27-07-2022 11:34:24
Nr: 2020
Ref : CHOL QC
Program Name: CHOL
OD.:0.2784
CONC: 115.179 mg/dL L
Normal Low: 150.0000
Normal High: 250.0000

RAJALAKSHMI LAB
Coralab Touch
27-07-2022 10:36:11
Nr: 2002
Ref : TGL QC
Program Name: TGL
OD.:0.1933
CONC: 160.781 mg/dL H
Normal Low: 68.0000
Normal High: 160.0000