



ACCURATE TESTING & CALIBRATION SERVICES

(Solution For All)

P. No. 111, K. No. 604, First Floor, Shyam Vihar, Part-II, Z-Block, Najafgarh, New Delhi-110043

E-mail : testingcalibrations@gmail.com | Web : www accuratetestingcalibration.com

M. : 9560785835, 7404479115, 8929915491



CC-3376

F. No : ATCS/QPM-32F REV.(01)26-02-2022

CALIBRATION CERTIFICATE

Page 1 of 1

ULR No. : CC337622000000993F	Certificate No. : ATCS22000000993
CUSTOMER NAME & ADDRESS:- M/s. Coronation Hospital Curzon Road, Dehradun 248001	Service Request No.&Date : 152 & 22-07-2022
	Date of Receipt : 22-07-2022
	Date of Calibration : 25-07-2022
	Sug. Cal. Due Date : 25-07-2023
	Issue Date : 26-07-2022

ITEM'S DESCRIPTION & IDENTIFICATION

Instrument Name : RPM Indicator fitted in Centrifuge	Range / Size : Up to 5250 RPM
Make : Remi	Least Count : 1 RPM
Model No. : R 8C	Visual Inspection : Ok
Serial No. : ZGHN-16737	Location : ICTC LAB
Party ID Mark : ----	

ENVIRONMENTAL CONDITIONS	Calibration Procedure : ATCS/CP/M-31
Temperature : (23 ± 10)°C	Reference Standard : IS: 1248
Relative Humidity : (50 ± 20)%	Calibration Performed at : Site

STANDARD EQUIPMENT USED			
Standard Instrument Name	Serial/Ident. No.	ULR/ Certificate No.	Valid upto
Tachometer	Q665619/ATCS/M/TM-01	C-210922-7-1	24-09-2022

CALIBRATION RESULT				
AVG. Nominal Value (UUC)	AVG. Observed Value (Master)	(±) Error	(±)Uncertainty	
(RPM)	(RPM)	(RPM)	(%)	
100	101.0	-1.0		
500	501.6	-1.6		
1000	1002	-2.0	1.0	
2500	2502	-2.0		
5000	5004	-4.0		

- NOTE :**
1. DUC = Device Under Calibration, UUC=Unit Under Calibration
 2. Expanded uncertainty of measurement at approx 95% confidence level for coverage factor (k) = 2
 3. Suggested next due date of calibration is given as requested by customer.
 4. DUC has been calibrated against lab standards are traceable to national & international standards.
 5. This result of calibration refer to only to the particular item submitted for calibration.
 6. Certificate issued is only scientific or industrial purposes only.
 7. The result produced in this certificate are valid under the stated conditions at the time of calibration.
 8. No report should be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the CEO, Accurate Testing & Calibration Services, New Delhi.

Vikram
Vikram Kumar
(Calibrated By)

Harshikesh Kumar
Harshikesh Kumar
(TM / QH)

Issued & Authorised By

Harshikesh Kumar

End of Report

