



ACCURATE TESTING & CALIBRATION SERVICES

(Solution For All)

P. No. 111, K. No. 604, First Floor, Shyam Vihar, Part-II, Z-Block, Najafgarh, New Delhi-110043
E-mail : testingcalibrations@gmail.com | Web : www accuratetestingcalibration.com
M. : 9560785835, 7404479115, 8929915491



CC-3376

F. No. : ATCS/QPM-32F REV (01)26-02-2022 **CALIBRATION CERTIFICATE**

Page 1 of 1

ULR No. : CC337622000000992F	Certificate No. : ATCS22000000992
CUSTOMER NAME & ADDRESS:- M/s. Coronation Hospital Curzon Road, Dehradun 248001	Service Request No.&Date : 152 & 22-07-2022
	Date of Receipt : 22-07-2022
	Date of Calibration : 22-07-2022
	Sug. Cal. Due Date : 22-07-2023
	Issue Date : 26-07-2022

ITEM'S DESCRIPTION & IDENTIFICATION

Instrument Name : Digital Thermometer	Range / Size : -50 to 300 °C
Make : Multi Thermometer	Least Count : 0.1 °C
Model No. : R-Tek	Visual Inspection : Ok
Serial No. : ----	Location : ICTC LAB
Party ID Mark : DTM-01	

ENVIRONMENTAL CONDITIONS	Calibration Procedure : ATCS/CP/TH-03
Temperature : (25 ± 03)°C	Reference Standard : ITS-90, NABL 129
Relative Humidity : (60 ± 15)%	Calibration Performed at : Lab

STANDARD EQUIPMENT USED			
Standard Instrument Name	Serial/Ident. No.	ULR/ Certificate No.	Valid upto
Dig. Thermometer With Sen. (4 Wire RTD)	ATCS/TH/DTM/001	TSC/21-22/7511-3	03-09-2022

CALIBRATION RESULT			
Standard Value (Master) (°C)	Observed Value (UUC) (°C)	(±) Error (°C)	(±)Uncertainty (°C)
-39.716	-40.0	-0.284	
0.015	0.0	-0.015	
50.065	50.0	-0.065	1.01
150.085	149.9	-0.185	
300.165	299.8	-0.365	

NOTE :

- DUC = Device Under Calibration, UUC=Unit Under Calibration
- Expanded uncertainty of measurement at approx 95% confidence level for coverage factor (k) = 2
- Suggested next due date of calibration is given as requested by customer.
- DUC has been calibrated against lab standards are traceable to national & international standards.
- This result of calibration refer to only to the particular item submitted for calibration.
- Certificate issued is only scientific or industrial purposes only.
- The result produced in this certificate are valid under the stated conditions at the time of calibration.
- No report should be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the CEO, Accurate Testing & Calibration Services, New Delhi.

Liyaquat Ali
(Calibrated By)



Issued & Authorised By

Harshikesh Kumar
(TM / QH)

End Of Report