

MK BEST CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

CERTIFICATE OF CALIBRATION

| FF/7.8/01 | | Page No | | | 1 of 1 |
|---|-------------------------------|--------------------------|--|-----------------------------------|------------|
| ULR No | CC334022000013010F | Date of Calibration | 21.07.2022 | Date of Receipt | 21.07.2022 |
| Certificate No | MKBL/22/07/0871-002 | Recom. Due Date | 20.07.2023 | Date of Issue | 25.07.2022 |
| CUSTOMER INFORMATION | | | DETAILS OF UNIT UNDER CALIBRATION | | |
| M/S ., SRI SUBHAM LAB , DEVANATHASWAMY NAGAR , EAST PONDY ROAD , VILLUPURAM - 605 103 | | | Description | MICROPIPETTE - 2 | |
| | | | Make / Model | --- | |
| | | | Range/Resolution | 5 to 50 µl | |
| | | | Serial No | VV100 | |
| | | | Identification No | SSV/EQP/CB/004 | |
| | | | Manufacturer Name | ERBA | |
| | | | Calibrated at | LAB | |
| STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards) | | | | | |
| S.No | Description | Id.No/SI. No | Certificate No | Validity | |
| 01 | Electronic Semi Micro Balance | MK/CAL-96/477904 | TVC SPL 21/07/863-01 | 23.07.2022 | |
| ENVIRONMENTAL & DUC CONDITIONS | | | REFERENCE STANDARD & ACCEPTANCE LIMIT | | |
| Temperature | 23 ± 1.5°C | Reference Std | ISO 8655 - 6 : 2002 | | |
| Humidity | 40 - 60 % RH | Procedure No | MKBCS - MBV - 03 | | |
| Condition of DUC Receipt | Good | | | | |
| CALIBRATION RESULTS | | | | | |
| I.VOLUME CALIBRATION | | | | | |
| S.No | DUC Reading (Mean) µl | STD Reading (Mean) µl | Deviation µl | Expanded Uncertainty (±) µl | |
| 1 | 10 | 10.15 | -0.15 | 1.18 | |
| 2 | 20 | 20.22 | -0.22 | | |
| 3 | 50 | 50.38 | -0.38 | | |
| Remarks : | | | | | |
| 1. The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2. | | | | | |
| 2. The Calibration Certificate Shall not be Reproduced Expect In Full,Without Written Approval Of The Laboratory. | | | | | |
| 3. The Recalibration Interval Should be Determined on the User Requirement. | | | | | |
| 4. The Results Stated In This Certificate Relate Only to the Item Calibrated. | | | | | |
| 5. The User Should Determine The Suitability Of The Instrument For Is Intended Use. | | | | | |
| 6. Resulted Volume Convert at 27°C of Water Temperature. | | | | | |
| 7. Expanded Uncertainty is also Included Correction Factors. | | | | | |
| x-x-x-x- End Of Certificate -x-x-x-x | | | | | |
| Calibrated by S.Murugesan (Calibration Engineer) | | | Authorised By L.Magesh (MD/QM) | | |

