

MK BEST CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

CERTIFICATE OF CALIBRATION

FF/7.8/01

Page No

1 of 1

ULR No	CC334022000013021F	Date of Calibration	21.07.2022	Date of Receipt	21.07.2022
Certificate No	MKBL/22/07/0874-001	Recom. Due Date	20.07.2023	Date of Issue	25.07.2022

CUSTOMER INFORMATION	DETAILS OF UNIT UNDER CALIBRATION	
M/S ., RAASI DIAGNOSTIC CENTRE , NO : 130 , TRICY ROAD , THURAIYUR - 621010 .	Description	MICROPIPETTE - 1
	Make / Model	FINN PIPETTE
	Range/Resolution	100 to 1000 µl
	Serial No	KW07722
	Identification No	RDC/EQP/CB/003
	Manufacturer Name	THERMO FISHER
	Calibrated at	LAB

STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)

S.No	Description	Id.No/Sl. No	Certificate No	Validity
01	Electronic Semi Micro Balance	MK/CAL-96/477904	TVC SPL 21/07/863-01	23.07.2022

ENVIRONMENTAL & DUC CONDITIONS REFERENCE STANDARD & ACCEPTANCE LIMIT

Temperature	23 ± 1.5°C	Reference Std	ISO 8655-6:2002
Humidity	40 - 60 % RH	Procedure No	MKBCS - MBV - 03
Condition of DUC Receipt	Good		

CALIBRATION RESULTS

1.VOLUME CALIBRATION

S.No	DUC Reading (Mean) µl	STD Reading (Mean) µl	Deviation µl	Expanded Uncertainty (±) µl
1	100	100.23	-0.23	7.29
2	200	200.35	-0.35	
3	400	400.43	-0.43	
4	800	800.56	-0.56	
5	1000	1000.73	-0.73	


Remarks :

1. The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.
2. The Calibration Certificate Shall not be Reproduced Expect In Full, Without Written Approval Of The Laboratory.
3. The Recalibration Interval Should be Determined on the User Requirement.
4. The Results Stated In This Certificate Relate Only to the Item Calibrated.
5. The User Should Determine The Suitability Of The Instrument For Its Intended Use.
6. Resulted Volume Convert at 27°C of Water Temperature.
7. Expanded Uncertainty is also Included Correction Factors.

x-x-x-x- End Of Certificate -x-x-x-x

Calibrated by

S. Murugesan
(Calibration Engineer)



Authorised By

L. Magesh
(MD/QM)