

HORIBA
Medical

HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2022-2023/1005835
04th Feb 2022


CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **ABX YUMIZEN H550** bearing serial number:
112YAXH03618 installed at **REDCLIFFE LIFETECH PVT. LTD., CHANDIGARH** as calibrated
on 04th Feb 2022

Calibrator : ABX MINOCAL
Lot No. : CX466
Expiry Date : 05th Feb 2022

The reports of Blank Cycle, Repeatability and Calibration Values were all found in
acceptable range.

Next calibration cycle is due on 03rd Aug 2022.



Shrish Dixit
(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**