

Calibration Certificate

Page No :- 1 of 1

Certificate No.	MACL/H/22/139-005	Calibration Certificate of CENTRIFUGE
Issue Date	07-07-2022	
Service Request No.	MACL/H/22/139	
Service Request Date	05-07-2022	
Calibration Date	05-07-2022	
Calibration Due Date (As per Customer request)	04-07-2023	

Name & Address of Customer	Identification on UUC (Unit Under Calibration)	
Dr. Vaibhav's Pathocity Path Lab & Imaging A-3/11, Basement Opposite HDFC Bank, Near Jwalaheri Maket, Paschim Vihar, New Delhi-110063.	Range/ Size	AS PER RESULT
	Make / Model No.	REMI/R-303
	Inst. I.D. No./Asset Code.	--
	Instrument Sr. No.	ZEDN-14064
	Least Count	AS PER RESULT
	Location	--
	Visual Inspection	O.K.

Calibration Procedure	MACL/M/CP- 80	Condition of Instrument	O.K.
Reference Standard	---	Calibration performed at	Lab
Environmental Condition	Temperature	20 ± 2 °C	Relative Humidity
			50 ± 20 % RH

Standard Equipmepment Used for Calibration				
Instrument Name	Make	I.D No./Sr.No.	Calibrated By	Certificate No.
1. Digital Tacho Meter	Kusam-Meco	MACL/TM-02	ACTL DELHI	ACTL/A/023/22-01
				Due Date
				21.01.2023

Calibration Results

Reading in rpm (UUC)	Reading in rpm (Ref.)
100	99.7
500	499.5
1000	998.2
2000	1999
3000	2998
3500	3497

Uncertainty of Measurement ± 1 %	The reported uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 correspond to confidence level of 95 %
----------------------------------	---

Notes:-
1. This Certificate refers only to the particular items submitted for calibration.
2. This Certificate shall not be reproduced, except in full, without the written permission of C.E.O. Micro Advance Calibration Lab , Ghaziabad.
3. Result Reported are Valid under the stated conditions of measurement.
4. Laboratory Standards are traceable to National Standards.
5. Calibration Certificate issue for weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scale/Tapes etc for scientific purpose only and should not be used for Trade / Commerical use.

CALIBRATED BY

Uma Shanker Sharma
Calibration Engineer



APPROVED BY

Mukesh Kumar Shakya
Authorized Signatory