



APEX QUALITY CERTIFICATION SERVICES PVT. LTD.

Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019
Ph.: 0141-4036776, 9414079483, 9636293333
Email: apex.qcs@gmail.com



CC-2298

CALIBRATION CERTIFICATE

ULR: CC229822000001378F	DISCIPLINE Mechanical	Page 01 of 01
CUSTOMER'S DETAILS	CERTIFICATE NO.:	AQCS/22071301.002
M/S RELIABLE DIAGNOSTIC CENTER (Chomu)	Date of Issue of Certificate	14/07/2022
Near Manglam City, Jaipur Road, Chomu Bypass Crossing,	Date of Calibration	13/07/2022
Radhaswami Bagh, Chomu (Rajasthan)	Recommended due date	12/07/2023

Service Request No.	220707/03	INSTRUMENT DESCRIPTION	
Service Request Date	07/07/2022	Nomenclature	Micropipette
DUC Receipt Date	12/07/2022	Make	Thermoscientific
DUC Condition	Ok	Model	FINNPIPETTE F2
Cal. Performed At	Lab	Sr. No. / Ident.	QW12656 / NA
Procedure Followed	CP/VM-01	Range	100-1000 μ l
Environmental Condition		Least Count	1 μ l
Temperature 25 \pm 2 $^{\circ}$ C	Rel. Hum. 50 \pm 10%	Location	--

STANDARD EQUIPMENT USED:

Nomenclature	Range	Uncertainty	Identification No.	Traceability Cert. No.	Valid up to
Weighing Balance	0 to 220 g	0.00002 & 0.0002 g	556300	TC/8479/2022	26-02-2023

CALIBRATION RESULT

DUC Value (μ l)	Observed Value on Standard (μ l)	Error (μ l)	Average Value at 27 $^{\circ}$ C (μ l)	Expanded Uncertainty \pm (μ l)
100	100.05	-0.05	100.05	0.66
250	250.09	-0.09	250.11	0.66
500	500.13	-0.13	500.17	0.66
750	750.19	-0.19	750.24	0.66
1000	1000.24	-0.24	1000.31	0.66

REMARKS :

- Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor k = 2 is as given above.
- The result reported in this certificate is valid at the time of and under the stated conditions of Measurement for the particular item and instrument identified above.
- This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur
- Calibration of volumetric measures done is meant for scientific and industrial purpose only, However, if used for commercial trading, additional recognition/approval shall be complied as required by Dept. of legal Calibration, regulatory bodies etc.
- Format Number.: F-06/00

Calibrated by

(V. B. Mathew)



Approved by

(R. R. Yadav)
Quality Manager