



APEX QUALITY CERTIFICATION SERVICES PVT. LTD.

Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019
Ph.: 0141-4036776, 9414079483, 9636293333
Email: apex.qcs@gmail.com



CALIBRATION CERTIFICATE

Page 01 of 01

CERTIFICATE NO.: AQCS/22071301.10003	Date of Calibration : 13/07/2022 Recommended due date : 12/07/2023
CUSTOMER'S DETAILS	INSTRUMENT DESCRIPTION
M/S RELIABLE DIAGNOSTIC CENTER (Chomu) Near Manglam City, Jaipur Road, Chomu Bypass Crossing, Radhaswami Bagh, Chomu (Rajasthan)	Nomenclature : Centrifuge Machine Make : Remi Model : R 303 Sr. No. : ZFIN-28618 Ident. No. : -- Range : 0-3500 RPM Least Count : As per results Location : -- Cal. Performed At : Site
Service Request No. : 220707/03 Date : 07/07/2022	

STANDARD EQUIPMENT USED:

Nomenclature	Identification No.	Traceability Certificate No.	Valid up to
Digital Tachometer	S101354	22000022012	08/06/2023

ENVIRONMENTAL CONDITIONS:

Temperature : $23 \pm 5^{\circ}\text{C}$ Relative Humidity : $50 \pm 10\%$

PROCEDURE FOLLOWED:

CP/RM-01

CALIBRATION RESULT

S. No.	Set position on DUC at	Measured Value on Standard in RPM
1	1	975
2	2	1496
3	3	2043
4	4	2558
5	5	3384

REMARKS :

- Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor $k = 2$ is $\pm 2.37\%$ of reading.
- The result reported in this certificate is valid at the time of and under the started conditions of Measurement for the particular item and instrument identified above.
- This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur
- Format Number.: F-06/00

Calibrated by

(V. B. Mathew)



Approved by

(R. R. Yadav)
Quality Manager