



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 2081

Distribution No.: 155-E

Month/Year: March/2022

Instrument ID: 903PES15247

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 29-04-2022[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	3.1	3	6.1	6.4	0.0320	-0.34	0.1	0.09	0.0050	0.15
RBC x10 <sup>6</sup> /µl	1	2.95	2.94	5.89	6.14	0.0070	-1.35	0.01	0.03	0.0020	-0.67
Hb g/dl	1	11.2	11.1	22.3	22.4	0.0210	-0.18	0.1	0.1	0.0070	0.00
HCT%	1	32.4	31.9	64.3	69.4	0.1290	-1.44	0.5	0.3	0.0230	0.67
MCV-fl	1	110	108	218	225.1	0.3410	-0.77	2	0.4	0.0310	2.70
MCH-Pg	1	38.2	37.7	75.9	73.1	0.0970	1.05	0.5	0.3	0.0220	0.67
MCHC-g/dl	1	35.2	34.4	69.6	64.65	0.1280	1.34	0.8	0.3	0.0150	1.69
Plt. x10 <sup>3</sup> /µl	1	159	154	313	351	1.21	-1.17	5	5	0.31	0.00
Retic %	2	8	7	15	15.5	0.26	-0.07	1	0.5	0.04	0.84

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=2 , Poly=49 L=05, E=0, Mono/Promono= , B1= P.M.=, Mye=43, Meta=03, Other=REMARKS - NEUTROPHILS -45 STABS 04
RBC Morphology	3	NORMOCHROMIC MILD ANISOCYTOSIS,FEW MICROCYTES NOTED ,PLATELETE APPEARS REDUCED IN NUMBER
Diagnosis	3	CHRONIC MYELOID LEUKEMIA (CML)
		Chronic Myeloid Leukemia (Chronic Phase)

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 155--E	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	320	318	89.94	90.25	5.66	3.77	4.4	5.98
<b>RBC x10<sup>6</sup>/µl</b>	1	320	320	88.13	87.81	6.56	3.13	5.31	9.06
<b>Hb g/dl</b>	1	320	320	85	91.25	5.63	3.13	9.37	5.62
<b>HCT%</b>	1	320	318	93.4	88.36	5.03	5.97	1.57	5.67
<b>MCV-fl</b>	1	320	317	94.01	95.58	4.42	1.58	1.57	2.84
<b>MCH-Pg</b>	1	320	317	90.22	88.01	4.1	5.36	5.68	6.63
<b>MCHC-g/dl</b>	1	320	318	93.08	90.25	4.09	4.72	2.83	5.03
<b>Plt. x10<sup>3</sup>/µl</b>	1	320	318	89.94	88.99	6.92	5.66	3.14	5.35
<b>ReticCount%</b>	2	320	298	90.94	88.26	6.04	6.38	3.02	5.36
<b>PS Assessment</b>	3	320	300	Satisfactory :87.16%, Borderline Sat. :6.89%, Unsatisfactory :5.95%					

**\*Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----