

**Calibration Certificate**

ADL/BO/MS/C/0297/21-22

This is to certify that **Mispa-i2 Specific Protein Analyzer** with Serial No: **21161631141** at **SRL Limited, Panjim** have been inspected and calibrated.

**Calibration Details**

Instrument Name	: Mispa-i2 Specific Protein Analyzer
Installation Date	: 15/02/2016
Calibration Date	: 04/Feb/2022
Calibration Certificate Due Date	: 03/Aug/2022
Checked the IC Card Reader	: Found OK
Checked Electronic Pipette Jack	: Found OK
Checked the Cuvette Motor	: Found OK
Checked Mother Board	: Found OK
Checked with Control	: Found OK

For Agappe Diagnostic Ltd.

Authorized Signatory



Name: *Bibin Baby*

Designation: *Sr. Engineer - Customer Service*

**AGAPPE DIAGNOSTICS LTD.** ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

**CORPORATE OFFICE / REAGENT PLANT**  
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.  
Tel: + 91 484 286 7000 | Email: agapped@agappe.in

**MUMBAI (REGISTERED OFFICE)**  
401 & 402, 4<sup>th</sup> Floor, Jaisingh Business Centre, 119,  
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.  
Tel: +91 22 4100 8000 | Email: mumbaioffice@agappe.in

**EQUIPMENT PLANT**  
X/588-CB, Block No. 32, KINFRA Small Industrial Park,  
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

**DELHI OFFICE**  
DSM 540, 5<sup>th</sup> Floor, DLF Tower, Shivaji Marg,  
New Delhi - 110 015, India.  
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

**KOLKATA OFFICE**  
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,  
Salt Lake City, Kolkata - 700 091.  
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

**BANGALORE OFFICE**  
5-6, II<sup>nd</sup> Floor, Red Cross Bhavan, No:26 Race Course Road,  
Bangalore - 560 001. Tel: +91 80 2228 8288  
Email: bangaloreoffice@agappe.in