

TECHNICAL RESOURCES

SALES, REPAIRS, AMC & FRANCHISE SERVICES OF ANALYTICAL, PRECISION, BIO-MEDICAL, BLOOD BANKING, IVF & HISTOPATHOLOGY LABORATORY EQUIPMENT.

A MSME Registered Unit, NSIC-CRISIL Rated SE 3C
A NABL Accredited Laboratory

Tel. : 022-28927149
Email : info@technical-resources.in
technicalresources@gmail.com
Handset : 9819103181
Webpage : www.technical-resources.in



SERVICE CENTRE : OFFICE NO. 2, TONES PARADISE, OPP. NEW PEDESTRAIN BRIDGE, NAVAGAON, DAHISAR (W), MUMBAI - 400 068.

Calibration Certificate

Calibration field: Thermal-Temperature	NABL Accreditation no: CC-2112
Accreditation valid upto: 11.04.2022	Certificate no: CC21122100000649
Requisition no: Sep2021-IW-239	Date of calibration: 15.09.2021
Calibration done at: At Lab	Agreed Next Calib On: 14.09.2022
Received On: 14.09.2021	Page no: 1 of 2

Customer details:	Reference used:
Name: SRL Diagnostic, Dr Avinash Phadke Lab	Master: Universal Calibrator UC12 with RTD Sensor
Address: Sumati, Ground Floor, Behind Saraswat Co-op-Bank, M.G Road, Panch Rasta, Mulund (W), Mumbai-400080	Master Ser. No. 20030512/202 ID NO: TM031/TM015
	Traceable Certificate M-201016-12-1/A
	Valid upto 20.10.2021

Details of DUC:	Condition of Item: Good
Equipment: Digital Thermometer	Range °C: -50 to 300
Make: ACETEQU	Accuracy ±°C : Not Specified
Model: ST-9283B	Resolution °C: 0.1
Serial No Not Specified	Length of sensor mm: 110
ID No: 1	Dia of sensor mm: 5
Location: Lab	Depth of sensor mm: 50

Ambient Temp(°C): 25.2	Humidity (%Rh): 56
-------------------------------	---------------------------

Remark:

1. Calibration procedure used: LAB SOP/04
2. The Reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor K=2, providing a level of confidence of approx. 95%.
3. The observations reported represent values at the time of measurement. They donot convey any long term stability information.
4. The standards maintained are traceable to National / International Standards through accredited laboratories.
5. Sticker affixed on the unit

Note:

1. This certificate refer only to particular instrument submitted for calibration.
2. This certificate shall not be reproduced.
3. Technical Resources shall not be responsible for any change in calibration date and performance specification on account of malfunctioning of the instrument due to damage caused to it after handing over it to the customer.


Authorised signatory
Sr Technical Manager
N Srinivas Rao

TECHNICAL RESOURCES

SALES, REPAIRS, AMC & FRANCHISE SERVICES OF AN ANALYTICAL
PRECISION, BIO-MEDICAL, BLOOD BANKING, IVF &
HISTOPATHOLOGY LABORATORY EQUIPMENT.

A MSME Registered Unit, NSIC-CRISIL Rated SE 3C
A NABL Accredited Laboratory

SERVICE CENTRE : OFFICE NO. 2, TONES PARADISE, OPP. NEW PEDESTRAIN BRIDGE, NAVAGAON, DAHISAR (W), MUMBAI - 400 068.

Tel : 022-28927149
Email : info@technical-resources.in
technicalresources@gmail.com
Handset : 9819103181
Webpage : www.technical-resources.in



Calibration data:

Equipment:	Digital Thermometer	Certificate/ULR no:	CC211221000000649
Serial no:	Not Specified	Date of calibration:	15.09.2021
ID No:	1	Page	2 of 2

RUN DETAILS

Set Temp. Display °C	Avg. Reading on DUC Display °C	Avg. Reading on Ref. Display °C	Deviation in (+/-) °C	Expanded Uncertainty in (+/-) °C
-20.0	-21.0	-20.04	0.94	0.600
5.0	6.4	5.15	-1.25	0.600
37.0	37.8	37.20	-0.62	0.600

END OF RESULT

21.09.2021
Certificate Issue Date:

Dilip
Checked By

Authorised Signatory
Sr Technical Manager
N Srinivas Rao

F-7.8.1.2