

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : Nihon Kohden Celltac Alpha

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	4.80	4.93	-0.42	-0.50
@ Total Leucocytes (WBC) count	/c.mm	4480.00	4,314.00	0.95	0.17
@ Haemoglobin	g/dL	15.40	15.68	-0.16	-0.03
@ MCV (Mean Corpuscular Volume)	fL	101.40	101.61	0.07	1.55
@ MCH (Mean Corpuscular Hb)	pg	31.80	31.69	0.00	0.89
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	31.30	30.90	0.07	-0.50
@ Platelet Count	1000/ μ L	98.00	105.95	-1.14	-1.68
@ Hematocrit	%	49.10	50.79	-0.25	0.27
@ RDW CV%	%	11.80	12.88	-1.00	2.78

Legend @ : z score \leq 2.0 - Acceptable
 | : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

DK Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohnoor Mall, Kiroli Road, Kuria (W),
 Mumbai - 400070

End of Report

MINU CLINICAL LABORATORY (558)
Outlier And Analyte Summary Report
Result Details For Cycle No 220104 and Sample No 02

Report Date : 23/05/2022

Instrument : Nihon Kohden Celltac Alpha

Analyte

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	5.60	5.56	-0.01	-0.50
@ Total Leucocytes (WBC) count	/c.mm	7720.00	6,388.00	0.84	0.17
@ Haemoglobin	g/dL	14.20	14.67	-0.59	-0.03
@ MCV (Mean Corpuscular Volume)	fL	90.20	92.34	-0.02	1.55
@ MCH (Mean Corpuscular Hb)	pg	25.30	25.54	-0.06	0.89
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	28.10	28.76	-0.31	-0.50
@ Platelet Count	1000/ μ L	343.00	319.44	0.19	-1.68
@ Hematocrit	%	50.70	51.29	-0.04	0.27
@ RDW CV%	%	15.20	15.54	-0.96	2.78

Legend @ : z score \leq 2.0 - Acceptable
| : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
Ⓢ : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

DK Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416
Commercial Building - 1A,
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070

End of Report

MINU CLINICAL LABORATORY (558)
 Outlier And Analyte Summary Report
 Result Details For Cycle No 210410 and Sample No 04

Report Date : 02/11/2021

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	40-70%	40-70%
@ Lymphocyte.	-	25-50%	25-50%
@ Eosinophil.	-	1-6%	1-6%

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	0
Evaluated Parameters	3
Outlier Parameters (X)	0
EQAS Score Haematology-Peripheral Smear Count	100.00 %

Qualitative

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Result Details For Cycle No 210410 and Sample No 01

Report Date : 02/11/2021

Instrument : Nihon Kohden Celltac Alpha

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	5.80	5.77	-0.10	0.02
@ Total Leucocytes (WBC) count	/c.mm	6480.00	6,390.00	0.21	0.68
@ Haemoglobin	g/dL	14.70	14.65	-0.11	-0.10
@ MCV (Mean Corpuscular Volume)	fL	87.70	85.05	0.01	-0.01
@ MCH (Mean Corpuscular Hb)	pg	25.20	25.30	-0.07	-0.16
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	28.80	29.85	-0.28	-0.14
@ Platelet Count	1000/ μ L	197.00	196.00	-0.68	-0.21
@ Hematocrit	%	51.10	49.20	-0.10	-0.15
@ RDW CV%	%	19.90	19.80	-1.28	-0.79

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable I : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

MINU CLINICAL LABORATORY (558)
Outlier And Analyte Summary Report
Result Details For Cycle No 210410 and Sample No 02

Report Date : 02/11/2021

Instrument : Nihon Kohden Celltac Alpha

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	4.60	4.56	0.13	0.02
@ Total Leucocytes (WBC) count	/c.mm	5290.00	4,795.00	1.14	0.68
@ Haemoglobin	g/dL	11.50	11.55	-0.09	-0.10
@ MCV (Mean Corpuscular Volume)	fL	86.40	84.15	-0.02	-0.01
@ MCH (Mean Corpuscular Hb)	pg	24.80	25.25	-0.24	-0.16
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	28.70	30.05	0.00	-0.14
@ Platelet Count	1000/ μ L	238.00	242.00	0.27	-0.21
@ Hematocrit	%	39.90	38.45	-0.20	-0.15
@ RDW CV%	%	19.90	19.70	-0.29	-0.79

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
⌚ : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416,
Commercial Building - 1A,
Kohinoor Mall, Kirod Road, Kuria (W),
Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable † : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____