

**MINU CLINICAL LABORATORY (558)**  
**Outlier And Analyte Summary Report**  
**Result Details For Cycle No 220104 and Sample No 01**

**Report Date : 23/05/2022**

**Instrument : -**

Analyte	Standard Unit	Result Value	Accepted Value
@ Crystal	-	Triple phosphate crystal	Triple phosphate crystal
@ Casts	-	No cast seen	No cast seen
# Other findings	-	Triple phosphate crystal	

Legend @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	1
Evaluated Parameters	2
Outlier Parameters (X)	0
<b>EQAS Score Clinical Path - Urine Sediment Analysis</b>	<b>100.00 %</b>

*Dr Nigam*

**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416,  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

End of Report



Metropolis Healthcare Ltd., EQAS

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	6,6.5
@ Specific Gravity	-	1.010	1.010,1.015,1.020
@ Glucose (urine)	-	2+	+1,+2,+3
@ Protein (Urine)	-	2+	+1,+2
@ Ketone (urine)	-	Negative	Negative
@ Blood (urine)	-	Negative	Negative
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
X Nitrite (urine)	-	Absent	Present

Legend @ : Acceptable  
X : Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Clinical Path - Urine Routine Chemistry	88.89 %

Dr Puneet Kumar Nigam  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416,  
Commercial Building - 1A,  
Door Mall, Kiroi Road, Kuria (W),  
Mumbai - 400070

End of Report

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Analyte	Instrument	Result Value	Standard Unit
X Nitrite (urine)	Manual	Absent	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: Error - Analytical  
 There is no sufficient evidence for Nitrite  
 and Nitrite is not reported for UTR in Meru  
 Clinical Lab

Corrective Action: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Reviewed by: 

Dated: 24/7/2021