



VECTOR BIOTEK PVT. LTD.

Factory & Registered Office: 425, NEW GIDC, KABILPORE, NAVSARI - 396 424. (GUJ.) INDIA
Toll Free No.: 1800-270-1027, Fax: +91-2637-265105, E-mail : vector@beaconindia.com
Web: www.vectorbiotekindia.com, CIN No. U33112GJ2002PTC04978

Calibration Certificate

Calibration of the **B AUTO 200** (Automatic Bio-chemistry Analyzer) has been carried out as per the recommendations of the principle company.

Customer Name : **Reliable Diagnostic centre, Chomu, Jaipur, Rajasthan.**

Date of Installation : **29/08/2020.**

Calibration Date : **08/06/2022.**

Next Calibration Due On : **07/12/2022.**

Tested By:

Checked By:



Performance Qualification

MODEL: B AUTO 200**Sr. No: AES2K0WH01023****Version: V4.8.3.0B3****Date: 08-06-2022**

PQ is performed at

Reliable Diagnostic centre, Chomu, Jaipur, Rajasthan.

According to the assay results that are comprised in this document the B AUTO 200 system can be approved for the routine operation.

System released for routine operation.

PERFORMANCE CALIBRATION			
LOT No: 26481			
TEST	TARGET	RANGE	RESULT
GLU	79.45	71.50-87.40	82.07
CHO	280.50	236-325	268.1
ALT	25.35	18.60-32.10	21.62
AST	40.90	35.90-45.90	36.71
ALP	166.30	135.30-197.30	197.2
BIL-D	0.32	0.20-0.44	0.38



VECTOR BIOTEK PVT. LTD.

Factory & Registered Office: 425, NEW GIDC, KABILPORE, NAVSARI - 396 424. (GUJ.) INDIA
Toll Free No.:1800-270-1027, Fax: +91-2637-265105, E-mail : vector@beaconindia.com
Web: www.vectorbiotekindia.com, CIN No. U33112GJ2002PTC04978

UA	4.40	3.70-5.10	4.72
ALB	4.25	3.50-5.00	3.7
UREA	32.20	24.00-40.40	31.13
TP	6.70	5.50-7.90	6.8
AMY	59.50	51.40-67.60	53.2
CA	9.70	7.90-11.50	9.4
CRE	1.85	1.30-2.40	1.47
BIL-T	1.20	0.90-1.50	1.04
TG	184.50	165-204	195.15

Note: Performance of the parameters are subject to traceability of Reagents, Manufacturer or Supplier.

Tested By:

Checked By:



VECTOR BIOTEK PVT. LTD.

Factory & Registered Office: 425, NEW GIDC, KABILPORE, NAVSARI - 396 424. (GUJ.) INDIA
Toll Free No.:1800-270-1027, Fax: +91-2637-265105, E-mail : vector@beaconindia.com
Web: www.vectorbiotekindia.com, CIN No. U33112GJ2002PTC04978

Operational Qualification

MODEL: B AUTO 200

Sr. No: AES2K0WH01023

Version: V4.8.3.0B3

Date: 08-06-2022

Operational Qualification Protocol run at Reliable Diagnostic centre, Chomu, Jaipur, Rajasthan.

#	Description	Status
1	Blank/Optics	OK
2	Assays(Parameter) Calibration	OK
3	Assay (Parameter) Control – BIORED	OK
4	Assays(Parameter) Samples(Patient)	OK
5	Programming and operational training completed	OK
6	Instrument operating training completed	OK

- Error msg. list in case of hardware issue: Refer Operational Manual page no. 75

I confirm that the above specified analyzer B AUTO 200 (Automatic Bio-chemistry Analyzer) is installed properly and its performance is within the specification.

Tested By:

Checked By: