



# ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



## CALIBRATION CERTIFICATE

Customer name And Address	M/S: Harley Pathlabs Pvt.Ltd. Basement A1, Sector-59, Noida-201301	ULR No.	CC223922000002606 F
		Service request No & date	P/02/06/06/2022
		Date of calibration	06/06/2022
		Valid up to	05/06/2023
		Date of issue	07/06/2022

Instrument Details			
Instrument name	Mini Centrifuge	I.D. / S. No.	CF-01
Make / Model No.	Remi / R-8C	Location	Lab
Range / Size	0 to 5250 rpm	Accuracy	-----
Least Count	10 rpm	Visual Inspection	OK

Detail of reference standards & Major equipments used			
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	-----/AC92792		
Certificate No.	QCTS/000229/10		
Calibration Validity	21/10/2022		
Calibration by	QCTS		

Environmental Condition	Temperature	20±2°C	Calibration Reference	IS: 12508
	Relative Humidity	50±10%	Calibration Procedure	CP-37

Calibration Results			
Serial No.	Standard value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	503.5	500	±0.2 %rdg.
02.	1002	1000	
03.	2006	2000	
04.	3006	3000	
05.	5009	5000	

Remarks:

- ◆ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ◆ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor = 2 such that the coverage probability corresponds to approximately 95% (3) The above results are valid at the time of and under the stated conditions measurement (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer

Calibrated By  
(Calibration Engg./TM)  
(MAHIPAT)  
Form No.-QP-47



7 / 10  
Appr.  
(QM/TM)  
(NEERAJYAGI)  
Page No. 1 of 1

---End of Report---