



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
ALL ANALYSER SUMMARY REPORT - JUNE 2022



Lab Name REDCLIFFE LIFETECH PVT LTD Lab No 15554
Constituent Group HbA1c Date of Result Entered 17/06/2022
PT item Lyophilized Whole blood Date of Report Published 07/07/2022

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SDI	U
						CV	SD			
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	129	7.02	4.34	0.30			0.05
			Beckman AU480/680/5800/DXC700AU	31	7.12	4.49	0.32			0.11
			Any Analyser	279	7.61	11.27	0.86	7.7 %	0.10	0.10

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:
REDCLIFFE LIFETECH PVT LTD
41/2, POCKET NO.2, VIVEK VIHAR, GMS ROAD,
DEHRADUN
UTTARAKHAND,248001

Coordinator Contact details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****