


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

|                                 |   |                        |                   |
|---------------------------------|---|------------------------|-------------------|
| <b>CERTIFICATE NO.</b>          | : 220711A/N-115   | <b>Page</b>            | : 1 of 1          |
| <b>CALIBRATED ON.</b>           | : 12 / 07 / 2022  | <b>Ref. Std.</b>       | : CG-11           |
| <b>CALIBRATION DUE.</b>         | : 11 / 01 / 2023  | <b>C.P. No.</b>        | : ACS/CP/07       |
| <b>SRF No. &amp; DATE</b>       | : 220711A/N Dt. 11.07.2022  | <b>Phy. Inspection</b> | : O.K.            |
| <b>ULR No. &amp; ISSUE DATE</b> | : CC342722000000081F Dt. 14.07.2022   | <b>Cal. At</b>         | : Site / Lab      |
| <b>CALIBRATED FOR</b>           | : M/s Redcliffe Lifetech Pvt. Ltd.<br>House no. 40-2, Vivek Vihar, Dehradun- 248001 |                        |                   |
| <b>ITEM CALIBRATED</b>          | : Digital Temp. Indicator of Refrigerator   |                        |                   |
|                                 | <b>Make</b> : Apteckdeals   | <b>Range</b>           | : (2 to 8) °C     |
|                                 | <b>L. Count</b> :   | <b>ID No.</b>          | : RCL/LAB/TEMP-03 |
|                                 | <b>Location</b> : Haematology Freezer   | <b>S. No.</b>          | :                 |

**ENVIRONMENT CONDITION:** Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

| STANDARD INSTRUMENTS USED : |                    |              |                 |            |
|-----------------------------|--------------------|--------------|-----------------|------------|
| INSTT. NAME                 | ID No. / S. No.    | TRACEABILITY | CERTIFICATE No. | DUE DATE   |
| Digital Calibrator          | ACS/UC/01 / 300561 | Anshaankan   | AIPL/2021/0784  | 11/11/2022 |

| RESULTS : ET Discipline   |                       |            |  |
|---------------------------|-----------------------|------------|--|
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) |  |
| 2.10                      | 2                     | 0.10       |  |
| 4.14                      | 4                     | 0.14       |  |
| 6.18                      | 6                     | 0.18       |  |
| 8.26                      | 8                     | 0.26       |  |
| (average of readings)     |                       |            |  |

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

Authn. Signatory  
S.K. Sharma  
CEO/TM