



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 220711A/N-117	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 12 / 07 / 2022	<b>Ref. Std.</b>	: CG-11
<b>CALIBRATION DUE.</b>	: 11 / 01 / 2023	<b>C.P. No.</b>	: ACS/CP/07
<b>SRF No. &amp; DATE</b>	: 220711A/N Dt. 11.07.2022	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342722000000083F Dt. 14.07.2022	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001		
<b>ITEM CALIBRATED</b>	: Digital Temp. Indicator of Refrigerator		
	<b>Make</b> : Apteckdeals	<b>Range</b>	: (2 to 8) °C
	<b>L. Count</b> :	<b>ID No.</b>	: RCL/LAB/TEMP-05
	<b>Location</b> : Sample Storage Fridge	<b>S. No.</b>	:
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C + 3°C } R.H. { 40% to 60% }		
<b>STANDARD INSTRUMENTS USED :</b>			
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784
			<b>DUE DATE</b>
			11/11/2022 -
<b>RESULTS</b>	: ET Discipline		
<b>* U. C. INSTT. READS (°C)</b>	<b>STANDARD READING (°C)</b>	<b>ERROR (°C)</b>	
2.08	2	0.08	
4.12	4	0.12	
6.20	6	0.20	
8.24	8	0.24	
	(average of readings)		

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is  $\pm 1.00^\circ\text{C}$  at coverage Factor  $k=2$ .

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

